

REF		Σ	SYSTEM
12017547122	12017547501	100	MODULAR ANALYTICS E170 cobas e 411 cobas e 601 cobas e 602

English

For use in the USA only

System information

For **cobas e 411** analyzer: test number 650

For MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers: Application Code Number 120

Intended use

Immunoassay for the in vitro quantitative determination of human insulin in human serum and plasma. The determination of insulin is utilized in the diagnosis and therapy of various disorders of carbohydrate metabolism, including diabetes mellitus and hypoglycemia.

The **electrochemiluminescence immunoassay** "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

Insulin is a peptide hormone with a molecular weight of approximately 6000 daltons. It is secreted by the B-cells of the pancreas and passes into circulation via the portal vein and the liver. Insulin is generally released in pulses, with the parallel glucose cycle normally about 2 minutes ahead of the insulin cycle.¹

The insulin molecule consists of two polypeptide chains, the α -chain with 21 and the β -chain with 30 amino acids. Biosynthesis of the hormone takes place in the β -cells of the islets of Langerhans in the form of single-chain proinsulin, which is immediately cleaved to give proinsulin. Specific proteases cleave proinsulin to insulin and C-peptide which pass into the bloodstream simultaneously. About half of the insulin, but virtually none of the C-peptide, is retained in the liver. Circulating insulin has a half-life of 3-5 minutes and is preferentially degraded in the liver, whereas inactivation or excretion of proinsulin and C-peptide mainly takes place in the kidneys.

The amino acid sequence of insulin has remained surprisingly constant during evolution, with the result that prior to the development of genetically engineered human insulin it was possible to successfully use porcine or bovine insulin in the therapy of diabetes mellitus.²

The action of insulin is mediated by specific receptors and primarily consists of facilitation of the uptake of sugar by the cells of the liver, fatty tissue and musculature; this is the basis of its hypoglycemic action.

Serum insulin determinations are mainly performed on patients with symptoms of hypoglycemia. They are used to ascertain the glucose/insulin quotients and for clarification of questions concerning insulin secretion, e.g. in the tolbutamide test and glucagon test or in the evaluation of oral glucose tolerance tests or hunger provocation tests.

Although the adequacy of pancreatic insulin synthesis is frequently assessed via the determination of C-peptide, it is still generally necessary to determine insulin. For example, therapeutic administration of insulins of non-human origin can lead to the formation of anti-insulin antibodies. In this case, measurement of the concentration of serum insulin shows the quantity of free - and hence biologically active - hormone, whereas the determination of C-peptide provides a measure of the patient's total endogenous insulin secretion.^{3,4,5}

A disorder in insulin metabolism leads to massive influencing of a number of metabolic processes. A too low concentration of free, biologically active insulin can lead to the development of diabetes mellitus. Possible causes of this include destruction of the β -cells (type I diabetes), reduced activity of the insulin or reduced pancreatic synthesis (type II), circulating antibodies to insulin, delayed release of insulin or the absence (or inadequacy) of insulin receptors.

On the other hand, autonomous, non-regulated insulin secretion is generally the cause of hypoglycemia. This condition is brought about by inhibition of gluconeogenesis, e.g. as a result of severe hepatic or renal failure, islet cell adenoma, or carcinoma. Hypoglycemia can, however, also be facilitated intentionally or unintentionally (factitious hypoglycemia).

In 3 % of persons with reduced glucose tolerance, the metabolic state deteriorates towards diabetes mellitus over a period of time. Reduced

glucose tolerance during pregnancy always requires treatment. The clearly elevated risk of mortality for the fetus necessitates intensive monitoring.

The Elecsys Insulin assay employs two monoclonal antibodies which together are specific for human insulin.

Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: Insulin from 20 μ L sample, a biotinylated monoclonal insulin-specific antibody, and a monoclonal insulin-specific antibody labeled with a ruthenium complex^{a)} form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex ($\text{Ru}(\text{bpy})_3^{2+}$)

Reagents - working solutions

The reagent rackpack is labeled as INSULIN.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL: Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-insulin-Ab~biotin (gray cap), 1 bottle, 10 mL: Biotinylated monoclonal anti-insulin antibody (mouse) 1 mg/L; MES^{b)} buffer 50 mmol/L, pH 6.0; preservative.
- R2 Anti-insulin-Ab~ $\text{Ru}(\text{bpy})_3^{2+}$ (black cap), 1 bottle, 10 mL: Monoclonal anti-insulin antibody (mouse) labeled with ruthenium complex 1.75 mg/L; MES buffer 50 mmol/L, pH 6.0; preservative.

b) MES = 2-morpholino-ethane sulfonic acid

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

2-methyl-2H-isothiazol-3-one hydrochloride

EUH 208 May produce an allergic reaction.

Product safety labeling follows EU GHS guidance.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Elecsys Insulin

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	4 weeks

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K₂-EDTA and K₃-EDTA plasma.

Criterion: Slope 0.9-1.1 + intercept within $\pm 0.8 \mu\text{U/mL}$ + coefficient of correlation ≥ 0.95 .

Stable for 4 hours at 20-25 °C, 2 days at 2-8 °C, 6 months at -20 °C (± 5 °C). Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Sample stability claims were established by experimental data by the manufacturer or based on reference literature and only for the temperatures/time frames as stated in the method sheet. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine specific stability criteria for its laboratory.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- [REF] 12017504122, Insulin CalSet, for 4 x 1.0 mL
- [REF] 11731416160, PreciControl Universal, for 4 x 3.0 mL or [REF] 05341787160, PreciControl Multimarker, for 6 x 2.0 mL

- General laboratory equipment
- MODULAR ANALYTICS E170 or **cobas e** analyzer

Additional materials for **cobas e** 411 analyzers:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, AssayCup, 60 x 60 reaction cups
- [REF] 11706799001, AssayTip, 30 x 120 pipette tips
- [REF] 11800507001, Clean-Liner

Additional materials for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use

- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change

- [REF] 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags

- [REF] 03023150001, WasteLiner, waste bags

- [REF] 03027651001, SysClean Adapter M

Additional material for all analyzers:

- [REF] 11298500160, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Calibration

Traceability: This method has been standardized using the 1st IRP WHO Reference Standard 66/304 (NIBSC).

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

Quality control

For quality control, use PreciControl Multimarker or PreciControl Universal.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Please note: Commercial controls may contain insulin of animal origin. When assessing results, the corresponding cross-reactivity of this test must be taken into account; see under "Analytical specificity".

Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in $\mu\text{U/mL}$ or pmol/L).

Conversion factors: $\mu\text{U/mL} \times 6.945 = \text{pmol/L}$
 $\text{pmol/L} \times 0.144 = \mu\text{U/mL}$

Limitations - interference

The effect of the following endogenous substances and pharmaceutical compounds on assay performance was tested. Interferences were tested up to the listed concentrations and no impact on results was observed.

Elecsys Insulin

Endogenous substances

Compound	Concentration tested
Bilirubin	≤ 1539 μmol/L or ≤ 90 mg/dL
Intralipid	≤ 1800 mg/dL
Biotin	≤ 246 nmol/L or ≤ 60 ng/mL
Rheumatoid factors	≤ 1200 IU/mL

Criterion: For concentrations of 0.2-2 μU/mL the deviation is ≤ 0.5 μU/mL. For concentrations > 2 μU/mL the deviation is ≤ 10 %.

Hemolysis interferes, as insulin-degrading peptidases are released from erythrocytes.⁶

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

There is no high-dose hook effect at insulin concentrations up to 20000 μU/mL or 138900 pmol/L.

In vitro tests were performed on 20 commonly used pharmaceuticals. No interference with the assay was found.

Samples from patients treated with bovine, porcine or human insulin sometimes contain anti-insulin antibodies which can affect the test results.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Measuring range

0.2-1000 μU/mL or 1.39-6945 pmol/L (defined by the lower detection limit and the maximum of the master curve). Values below the lower detection limit are reported as < 0.2 μU/mL (< 1.39 pmol/L). Values above the measuring range are reported as > 1000 μU/mL (> 6945 pmol/L).

Lower limits of measurement

Lower detection limit of the test

Lower detection limit: 0.2 μU/mL (1.39 pmol/L)

The Lower Detection Limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the lowest standard (master calibrator, standard 1 + 2 SD, repeatability study, n = 21).

Dilution

Not necessary due to the broad measuring range.

Expected values

Studies with the Elecsys Insulin assay conducted in a clinical center in Germany with samples from 57 healthy, fasting individuals gave the following results (5th-95th percentile range):

2.6-24.9 μU/mL (17.8-173 pmol/L)

Status: Elecsys Insulin MCE, study No.: B99P027 of 29 March 2001.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents and pooled human sera in a modified protocol (EP5-A) of the CLSI (Clinical and Laboratory Standards Institute): 6 times daily for 10 days (n = 60); repeatability on MODULAR ANALYTICS E170 analyzer, n = 21. The following results were obtained:

cobas e 411 analyzer					
			Repeatability		
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
Human serum 1	6.36	44.2	0.122	0.847	1.9
Human serum 2	20.9	145	0.391	2.71	1.9
Human serum 3	747	5188	15.1	105	2.0

cobas e 411 analyzer					
			Intermediate precision		
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
Human serum 1	6.36	44.2	0.163	1.11	2.6
Human serum 2	20.9	145	0.593	4.10	2.8
Human serum 3	747	5188	18.6	129	2.5

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers					
	Repeatability				
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
Human serum 1	5.93	41.2	0.09	0.62	1.5
Human serum 2	14.5	101	0.13	0.92	0.9
Human serum 3	49.9	346	0.58	4.05	1.2
Human serum 4	399	2768	3.32	23.1	0.8

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers					
	Intermediate precision				
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
Human serum 1	6.85	47.6	0.336	2.33	4.9
Human serum 2	16.7	116	0.616	4.28	3.7
Human serum 3	55.1	383	1.86	12.9	3.4
Human serum 4	425	2949	10.0	69.6	2.4

Precision was determined using Elecsys reagents and controls in a protocol (EP5-A2) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplicate each for 21 days (n = 84). The following results were obtained:

cobas e 411 analyzer					
			Repeatability		
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
PreciControl MM ^{c)} 1	23.7	165	0.270	1.88	1.1
PreciControl MM2	81.7	567	1.14	7.92	1.4

c) MM = Multimarker

cobas e 411 analyzer					
			Intermediate precision		
Sample	Mean		SD		CV
	μU/mL	pmol/L	μU/mL	pmol/L	%
PreciControl MM1	23.7	165	0.834	5.79	3.5
PreciControl MM2	81.7	567	3.04	21.1	3.7

Elecsys Insulin



MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers					
Sample	Mean		Repeatability		CV
	μU/mL	pmol/L	μU/mL	pmol/L	
PreciControl MM1	21.9	152	0.712	4.94	3.2
PreciControl MM2	74.3	516	2.72	18.9	3.7

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers					
Sample	Mean		Intermediate precision		CV
	μU/mL	pmol/L	μU/mL	pmol/L	
PreciControl MM1	21.9	152	0.926	6.43	4.2
PreciControl MM2	74.3	516	3.42	23.8	4.6

Method comparison

a) A comparison of the Elecsys Insulin assay (y) with the Enzymun-Test Insulin method (x) using clinical samples gave the following correlations (μU/mL):

Number of samples measured: 99

Passing/Bablok⁷ Linear regression

$y = 1.00x - 1.16$ $y = 0.92x + 0.59$

$r = 0.844$ $r = 0.958$

The sample concentrations were between 3.9 and 80 μU/mL (27 and 550 pmol/L).

b) A comparison of the Elecsys Insulin assay (y) with a commercially available Insulin test (x) using clinical samples gave the following correlations (μU/mL):

Number of samples measured: 99

Passing/Bablok⁷ Linear regression

$y = 0.89x - 0.62$ $y = 0.93x - 1.02$

$r = 0.935$ $r = 0.981$

The sample concentrations were between 1 and 118 μU/mL (7 and 820 pmol/L).

Analytical specificity

For the monoclonal antibodies used, the following cross-reactivities were found:

	Concentration tested	Cross-reactivity %
Bovine insulin	17360 pmol/L	25.0
Porcine insulin	8334 pmol/L	19.2
Human proinsulin	1000 ng/mL	0.05
C-peptide	100 ng/mL	n.d. ^{d)}
Glucagon	1000 pg/mL	n.d.
Somatostatin	100 pg/mL	n.d.
Insulin-like growth factor I	6579 pmol/L	0.04

d) n.d. = not detectable

Results for cross-reactivity with recombinant insulin analogs in a number of insulin methods have been published for example by two groups in France and the USA.^{8,9,10} The following results were published by Owen et al.⁹ for the Elecsys Insulin assay:

Insulin lispro, insulin aspart, and insulin glargine were each tested in concentrations of 30, 100, 300, and 1000 mIU/L in the absence of insulin. The results obtained were below the detection limit of the Elecsys Insulin assay (< 0.2 μU/mL or < 1.39 pmol/L) at all the concentrations tested.

Moreover, these results also correlate with those published earlier by Sapin et al. for insulin lispro.⁸

References

- Lang DA, Matthews DR, Peto J, et al. Cyclic oscillations of basal plasma glucose and insulin concentrations in human beings. *N Engl J Med* 1979;301:1023-1027.
- Fiedler H. Basiswissen Labordiagnostik: Diabetes mellitus und metabolisches Syndrom. Broschüre Roche Diagnostics 1999;14,67 Best.-Nr. 1951769.
- Arnqvist H, Olsson PO, von Schenck H. Free and Total Insulin as Determined after Precipitation with Polyethylene Glycol: Analytical Characteristics and Effects of Sample Handling and Storage. *Clin Chem* 1987;33(1):93-96.
- Gerbitz KD. Pankreatische B-Zellen Peptide: Kinetik und Konzentration von Proinsulin, Insulin und C-Peptid in Plasma und Urin, Probleme der Meßmethoden, klinische Aussage und Literaturübersicht. *J Clin Chem Clin Biochem* 1980;18(6):313-326.
- Clark PM. Assays for insulin, proinsulin(s) and C-peptide. *Ann Clin Biochem* 1999;36(5):541-564.
- Chevenne D, Letailleur A, Trivin F, et al. Effect of Hemolysis on the Concentration of Insulin in Serum Determined by RIA and IRMA. *Clin Chem* 1998;44(2):354-356.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.
- Sapin R, Le Galudec V, Gasser F, et al. Elecsys Insulin Assay: Free Insulin Determination and the Absence of Cross-Reactivity with Insulin Lispro. *Clin Chem* 2001;47:602-605.
- Owen WE, Roberts WL. Letter to the Editor: Cross-Reactivity of Three Recombinant Insulin Analogs with Five Commercial Insulin Immunoassays. *Clin Chem* 2004;50(1):257-259.
- Sapin R. Review: Insulin Assays: Previously Known and New Analytical Features. *Clin Lab* 2003;49(3-4):113-121.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume after reconstitution or mixing
GTIN	Global Trade Item Number

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2020, Roche Diagnostics

Elecsys Insulin

cobas[®]



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com



Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336