

Elecsys PTH (1-84)

cobas®

REF



SYSTEM

05608546190

100

cobas e 411

cobas e 601

cobas e 602

English

System information

For **cobas e 411** analyzer: test number 930

For **cobas e 601** and **cobas e 602** analyzers: Application Code Number 273

Please note

The measured PTH (1-84) value of a patient's sample can vary depending on the testing procedure used. The laboratory finding must therefore always contain a statement on the PTH (1-84) assay method used. PTH (1-84) values determined on patient samples by different testing procedures cannot be directly compared with one another and could be the cause of erroneous medical interpretations. Therefore, the results reported by the laboratory to the physician should include: "The following results were obtained with the Elecsys PTH (1-84) assay. Results from assays of other manufacturers cannot be used interchangeably."

The performance characteristics for this assay have not been established for pediatric specimens.

Intended use

Immunoassay for the in vitro quantitative determination of bioactive parathyroid hormone, PTH (1-84) in human serum and plasma for the differential diagnosis of hypercalcemia and hypocalcemia. This assay can be used intraoperatively.

The **electrochemoluminescence immunoassay "ECLIA"** is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

Parathyroid hormone (PTH) 1-84 is the full-length single-chain 84-amino-acid peptide produced by the parathyroid glands in response to decreased extracellular concentrations of ionized calcium. Its main role is to increase serum calcium levels by stimulating the release of calcium from bone and its renal re-absorption in the distal tubule. In the proximal tubule, PTH stimulates the synthesis of calcitriol which in turn increases intestinal absorption of calcium and exerts an endocrine feed-back on the secretion of PTH at the parathyroid level. PTH also decreases the renal re-absorption of phosphate in the proximal tubule, thereby decreasing serum phosphate.¹ PTH (1-84) has a half-life of only a few minutes and is cleaved into various fragments and cleared very rapidly from the circulation.²

Parathyroid gland disorders lead to elevated or depressed blood calcium levels (hypercalcemia or hypocalcemia) caused by a change in the secretion of PTH. Detection of subfunctioning parathyroid glands (hypoparathyroidism) requires the use of a highly sensitive test in order to be able to measure PTH levels well below normal.³

Hyperfunctioning of the parathyroid glands results in an increased secretion of PTH (hyperparathyroidism). Primary causes are adenomas of the parathyroid glands. In secondary hyperparathyroidism the blood calcium level is low as a result of other pathological states (e.g. vitamin D deficiency).³

PTH measurement is routinely prescribed in patients with chronic renal failure to identify renal osteodystrophy subtypes and adapt treatment, and in non-renal patients to explore any disorder of calcium-phosphate metabolism.⁴

Measurement of PTH (1-84) in patients with chronic kidney disease (CKD)

PTH levels serve as surrogates for bone histology in patients with end-stage renal disease (ESRD), and they are an essential guide to ongoing clinical management, particularly during the treatment of secondary hyperparathyroidism with vitamin D sterols.⁵

The Kidney Disease Outcomes Quality Initiative (KDOQI) and Kidney Disease Improving Global Outcomes (KDIGO) guidelines recommend that serum PTH concentration should be measured regularly in patients with CKD beginning with CKD stage 3 (stage 2 for children).⁶ When the PTH concentration is above the target values, drugs that decrease PTH

secretion such as active vitamin D compounds or calcimimetic agents may be given and the doses are then adapted according to the evolution of the PTH concentration. By contrast, if the PTH concentration is below the target range, any treatment that may decrease PTH secretion is stopped to avoid adynamic bone disease and associated extra-skeletal calcifications.⁷ Low PTH levels induced by high dialysate calcium levels were found to be associated with an increased risk of cardiovascular death.⁸

The Elecsys PTH (1-84) assay is a third generation PTH assay⁴ as it specifically measures the biologically intact molecule of PTH, PTH (1-84). This can be of advantage in chronic renal failure patients as it was shown that fragments of PTH (i.e. PTH 7-84) accumulate in patients on dialysis presumably due to reduced excretion.⁹ However, the proportion of these fragments increases as glomerular filtration rate (GFR) falls, suggesting that this relationship might vary with the severity of renal failure.¹⁰

Since then, several studies have demonstrated that, in animal or cellular models, PTH (7-84) exerts effects that are opposite to those of PTH (1-84) (decrease in serum calcium and phosphate, inhibition of bone resorption), and is produced by the parathyroid glands in response to an increase in serum calcium levels.¹¹

Also the ratio between PTH (1-84) and PTH (7-84) was proposed as a predictor for the severity of secondary hyperparathyroidism in dialysis patients.¹²

Clinical investigations in intraoperative use

The determination of PTH intraoperatively during adenoma resection in the parathyroid glands has been reported for primary hyperparathyroidism,^{13,14} secondary hyperparathyroidism relating to renal failure,^{15,16} and tertiary hyperparathyroidism post renal transplant surgery.¹⁷

Because PTH has a reported half-life of 3-5 minutes,¹⁸ a significant drop in PTH levels after resection of the abnormal gland enables the surgeon to assess the completeness of resection and whether all hyperfunctioning parathyroid tissue has been removed from the patient.¹⁹

The National Academy of Clinical Biochemistry recommended routine use of intraoperative PTH testing for patients undergoing surgery for primary hyperparathyroidism, both in initial surgeries and in reoperative procedures.²⁰

Selective measurement of the biologically intact PTH (1-84) molecule may be advantageous during parathyroid surgery, as there is evidence that PTH (1-84) decreases more rapidly than "non-1-84" PTH immediately after parathyroidectomy.⁴

The Elecsys PTH (1-84) assay employs a sandwich test principle in which a biotinylated monoclonal antibody reacts with the N-terminal fragment PTH and a monoclonal antibody labeled with a ruthenium complex^{a)} reacts with the C-terminal fragment PTH.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 50 µL of sample, a biotinylated monoclonal PTH-specific antibody, and monoclonal PTH-specific antibody labeled with a ruthenium complex form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

Reagents - working solutions

The reagent rackpack is labeled as PTH (1-84).

Elecsys PTH (1-84)

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-PTH-Ab~biotin (gray cap), 1 bottle, 7 mL:
Biotinylated monoclonal anti-PTH antibody (mouse) 2.0 mg/L;
phosphate buffer 100 mmol/L, pH 7.0; preservative.
- R2 Anti-PTH-Ab~Ru(bpy)₃²⁺ (black cap), 1 bottle, 7 mL:
Monoclonal anti-PTH antibody (mouse) labeled with ruthenium
complex 1.0 mg/L; phosphate buffer 100 mmol/L, pH 7.0;
preservative.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

2-methyl-2H-isothiazol-3-one hydrochloride

EUH 208 May produce an allergic reaction.

Product safety labeling follows EU GHS guidance.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	84 days
on the analyzers	56 days

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

K₂-EDTA, K₃-EDTA and Li-heparin plasma as well as Li-heparin plasma tubes containing separating gel.

Because of the short half-life of PTH, it is recommended that, when serum is needed, the blood must be centrifuged immediately.

Preference should be given to K₂-EDTA or K₃-EDTA plasma, as it is stable longer than serum.

Criterion: Method comparison serum versus plasma, slope 0.9-1.1 + intercept within $\leq \pm 2 \times$ Limit of Blank + coefficient of correlation ≥ 0.95 .

Serum: Stable for 7 hours at 15-25 °C, 24 hours at 2-8 °C, 6 months at -20 °C (± 5 °C). Freeze only once.

Plasma: Stable for 24 hours at 15-25 °C, 48 hours at 2-8 °C, 6 months at -20 °C (± 5 °C). Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- REF 05608554190, PTH (1-84) CalSet, for 4 x 1.0 mL
- REF 05618860190, PreciControl Varia, for 4 x 3.0 mL
- REF 11732277122, Diluent Universal, 2 x 16 mL sample diluent or
REF 03183971122, Diluent Universal, 2 x 36 mL sample diluent

General laboratory equipment

cobas e analyzer

Additional materials for the **cobas e** 411 analyzer:

- REF 11662988122, ProCell, 6 x 380 mL system buffer
- REF 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- REF 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- REF 11933159001, Adapter for SysClean
- REF 11706802001, AssayCup, 60 x 60 reaction cups
- REF 11706799001, AssayTip, 30 x 120 pipette tips
- REF 11800507001, Clean-Liner

Additional materials for **cobas e** 601 and **cobas e** 602 analyzers:

- REF 04880340190, ProCell M, 2 x 2 L system buffer
- REF 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- REF 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- REF 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- REF 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- REF 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags
- REF 03023150001, WasteLiner, waste bags
- REF 03027651001, SysClean Adapter M

Additional materials for all analyzers:

- REF 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

cobas e 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Calibration

Traceability: This method has been standardized against the WHO international standard 95/646.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Elecsys PTH (1-84)

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 12 weeks when using the same reagent lot
- after 7 days when using the same reagent kit on the analyzer
- as required: e.g. quality control findings outside the defined limits

Quality control

For quality control, use PreciControl Varia.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in pg/mL or pmol/L).

Conversion factors: $\text{pg/mL} \times 0.106 = \text{pmol/L}$
 $\text{pmol/L} \times 9.43 = \text{pg/mL}$

Limitations - interference

The assay is unaffected by icterus (bilirubin $\leq 1112 \mu\text{mol/L}$ or $\leq 65 \text{ mg/dL}$), lipemia (Intralipid $\leq 1500 \text{ mg/dL}$), IgG $\leq 28 \text{ g/L}$, IgM $\leq 8 \text{ g/L}$, IgA $\leq 16 \text{ g/L}$, albumin $\leq 120 \text{ g/L}$ and biotin ($\leq 205 \text{ nmol/L}$ or $\leq 50 \text{ ng/mL}$).

The assay is affected by hemolysis $\geq 0.1 \text{ g/dL}$. Do not analyze samples that show visible signs of hemolysis.

Criterion: Recovery within $\pm 12 \%$ of initial value for samples $\geq 25 \text{ pg/mL}$ or $\pm 3 \text{ pg/mL}$ for samples $< 25 \text{ pg/mL}$.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. $> 5 \text{ mg/day}$) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1500 IU/mL .

There is no high-dose hook effect at PTH (1-84) concentrations up to 30000 pg/mL (3180 pmol/L).

In vitro tests were performed on 18 commonly used and 5 special pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Measuring range

$5.50\text{--}2300 \text{ pg/mL}$ or $0.583\text{--}244 \text{ pmol/L}$ (defined by the Limit of Detection and the maximum of the master curve). Values below the Limit of Detection are reported as $< 5.50 \text{ pg/mL}$ ($< 0.583 \text{ pmol/L}$). Values above the measuring range are reported as $> 2300 \text{ pg/mL}$ ($> 244 \text{ pmol/L}$).

Lower limits of measurement

Limit of Blank, Limit of Detection and Limit of Quantitation

Limit of Blank = 3.50 pg/mL

Limit of Detection = 5.50 pg/mL

Limit of Quantitation = 10.0 pg/mL

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A requirements.

The Limit of Blank is the 95th percentile value from $n \geq 60$ measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is defined as the lowest amount of analyte in a sample that can be accurately quantitated with a total allowable relative error of $\leq 30 \%$.

Dilution

Samples with PTH (1-84) concentrations above the measuring range can be diluted with Diluent Universal. The recommended dilution is 1:2 (either automatically by the analyzers, or manually). The concentration of the diluted sample must be $\geq 1150 \text{ pg/mL}$ ($\geq 122 \text{ pmol/L}$).

After manual dilution, multiply the result by the dilution factor.

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

Expected values

Normal values in apparently healthy individuals

The normal value range was determined in a clinical study using 596 samples from apparently healthy individuals. The reference population was selected according to normal clinical chemistry parameters, normal hematology results, no vitamin D intake and normal calcium values as determined by flame photometry. The values given are only indicative and may vary from other published data.

N = 596	PTH (1-84) range	
	pg/mL	pmol/L
Mean	31.3	3.32
2.5 th percentile	14.9	1.58
97.5 th percentile	56.9	6.03

The subgrouping of the above cohort according to the vitamin D (25-OH) level shows the inverse relationship between the concentrations of PTH (intact) and PTH (1-84) and vitamin D (25-OH).

N	Vitamin D (25-OH)	PTH (intact) median		PTH (1-84) median	
	ng/mL	pg/mL	pmol/L	pg/mL	pmol/L
339	≤ 20	41.4	4.39	32.2	3.41
157	> 20 and < 30	37.0	3.92	29.0	3.07
100	≥ 30	33.0	3.50	24.9	2.64

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, pooled human sera and controls in a protocol (EP5-A2) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplicate each for 21 days ($n = 84$). The following results were obtained:

cobas e 411 analyzer								
		Repeatability				Intermediate precision		
Sample	Mean		SD		CV	SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L	%
HS* 1	339	35.9	5.29	0.561	1.6	13.9	1.47	4.1
HS 2	1398	148	19.5	2.07	1.4	43.4	4.60	3.1

Elecsys PTH (1-84)

cobas e 411 analyzer							
		Repeatability			Intermediate precision		
Sample	Mean		SD		CV	SD	
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L
HS 3	13.2	1.40	0.976	0.104	7.4	1.23	0.130
HS 4	27.6	2.93	0.870	0.092	3.2	1.28	0.136
PC** Varia 1	35.1	3.72	1.42	0.151	4.1	2.04	0.216
PC Varia 2	103	10.9	4.07	0.431	3.9	4.96	0.526

* HS = Human serum

** PC = PreciControl

cobas e 601 and cobas e 602 analyzers							
		Repeatability			Intermediate precision		
Sample	Mean		SD		CV	SD	
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L
HS 1	6.35	0.673	0.221	0.023	3.5	0.394	0.042
HS 2	38.3	4.06	0.450	0.048	1.2	1.16	0.123
HS 3	339	35.9	2.58	0.274	0.8	9.09	0.964
HS 4	2028	215	14.9	1.58	0.7	60.3	6.39
PC Varia 1	41.2	4.37	0.562	0.060	1.4	0.656	0.070
PC Varia 2	118	12.5	1.84	0.195	1.6	2.09	0.222

Method comparison

A comparison of the Elecsys PTH (1-84) assay (y) with the Elecsys PTH assay (x) using clinical samples gave the following correlations (pg/mL):

Number of samples measured: 1347

Passing/Bablok²¹ Linear regression
 $y = 0.668x + 3.02$ $y = 0.555x + 10.2$
 $r = 0.927$ $r = 0.987$

The sample concentrations were between 9.38 and 2514 pg/mL (0.994 and 266 pmol/L).

Analytical specificity

- ≤ 0.1 % cross-reactivities: Osteocalcin, β-CrossLaps (collagen-fragment), and bone-specific alkaline phosphatase
- ≤ 0.1 % cross-reactivities: PTH (1-34), PTH (7-84)
- For the N-terminal PTH related peptide (PTH-RP) no cross-reactivity was found in an epitope scan with the monoclonal antibody reactive with PTH N-terminal fragment used in this assay.

References

- Thomas L. Parathyroid hormone (PTH). Clinical Laboratory Diagnosis. TH-Books, Frankfurt. 1st english edition 1998: 248-250.
- Segre GV, Niall HD, Habener JF, et al. Metabolism of parathyroid hormone: physiological and clinical significance. Am J Med 1974;56:774-784.
- Nussbaum S, Potts JT. Advances in Immunoassays for Parathyroid Hormone. Clinical Applications to Skeletal Disorders of Bone and Mineral Metabolism. In Bilezikian JP, Levine MA, Marcus R (eds). The Parathyroids: Basic and Clinical Concepts. Raven Press, New York 1994:157-169.
- Souberbielle JC, Friedlander G, Cormier C. Practical considerations in PTH testing. Clin Chim Acta 2006;366:81-89.
- Goodman WG, Salusky IB, Juppner H. New lessons from old assays: parathyroid hormone (PTH), its receptors, and the potential biological relevance of PTH fragments. Nephrol Dial Transplant 2002;17:1731-1736.
- Ketteler M, Grahame JE, Evenepoel P, et al. Revisiting KDIGO clinical practice guideline on chronic kidney disease-mineral and bone disorder: a commentary from a Kidney Disease: Improving Global Outcomes controversies conference. International Society of Nephrology 2015 Meeting report. Kidney Int 2015;87:502-508.

- Souberbielle JC, Roth H, Fouque D. Parathyroid hormone measurement in CKD. Kidney Int 2010;77:93-100.
- Merle E, Roth H, London GM, et al. Low parathyroid hormone status induced by high dialysate calcium is an independent risk factor for cardiovascular death in hemodialysis patients. Kidney Int 2016;89(3):666-674.
- Waller S, Ridout, Cantor T, et al. Differences between "intact" PTH and 1-84 PTH assays in chronic renal failure and dialysis. Pediatr Nephrol 2005;20:197-199.
- Brossard JH, Lepage R, Cardinal H, et al. Influence of glomerular filtration rate on non-(1-84) parathyroid hormone (PTH) detected by intact PTH assays. Clin Chem 2000;46:697-703.
- Divieti P, John MR, Juppner H, et al. Human PTH-(7-84) inhibits bone resorption in vitro via actions independent of the type 1 PTH/PTHrP receptor. Endocrinology 2002;143:171-176.
- Tanaka M, Komaba H, Itoh K, et al. The whole-PTH/intact-PTH ratio is a useful predictor of severity of secondary hyperparathyroidism. NDT Plus 2008;1:[Suppl 3]:iii59-iii62.
- Bergenfels A, Nordén NE, Ahrén B. Intraoperative fall in plasma levels of intact parathyroid hormone after removal of one enlarged parathyroid gland in hyperthyroid patients. Eur J Surg 1991;157:109-112.
- Ohe MN, Santos RO, Kunii IS, et al. Usefulness of a rapid immunometric assay for intra-operative parathyroid hormone measurements. Braz J Med Biol Res 2003;36(6):715-721.
- Seehofer D, Rayes N, Ulrich F, et al. Intra-operative measurement of intact parathyroid hormone in renal hyperparathyroidism by an inexpensive routine assay. Langenbecks Arch Surg 2001;386(6):440-443.
- Seehofer D, Rayes N, Klupp J, et al. Predictive value of intact parathyroid hormone measurement during surgery for renal hyperparathyroidism. Langenbecks Arch Surg 2005;390(3):222-229.
- Haustein SV, Mack E, Starling JR, et al. The role of intra-operative parathyroid hormone testing in patients with tertiary hyperparathyroidism after renal transplantation. Surgery 2005;138(6):1066-1071.
- Maier GW, Kreis ME, Renn W, et al. Parathyroid hormone after adenectomy for primary hyperparathyroidism: A study of peptide hormone elimination kinetics in humans. Jour Clin Endocrinol Metab 1998;83(11):3853-3856.
- Carter AB, Howanitz TJ. Intra-operative testing for parathyroid hormone: a comprehensive review of the use of the assay and the relevant literature. Arch Pathol Lab Med 2003;127:1424-1442.
- Nichols JH, Christenson RH, Clarke W, et al. National Academy of Clinical Biochemistry Laboratory Medicine Practice Guidelines: Evidence Based Practice for Point of Care Testing. AACCPress:2007.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume after reconstitution or mixing

Elecsys PTH (1-84)

cobas[®]

GTIN

Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2019, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

