

DISCOVERY ChromoMap DAB Kit

REF

760-159

05266645001



INTENDED USE

For Research Use Only. Not for use in diagnostic procedures.

SUMMARY AND EXPLANATION

Automated immunohistochemistry (IHC) is a powerful tool for demonstrating the presence of unique antigens in tissues and cells. The ChromoMap DAB Kit was designed for chromogenic antigen localization in frozen sections or formalin-fixed, paraffin-embedded (FFPE) tissue sections, to be used in conjunction with the Discovery series of instruments and Ventana Medical Systems' conjugates and ancillary reagents for optimal performance.

MATERIAL PROVIDED

DISCOVERY ChromoMap DAB Kit (RUO) contains sufficient reagent for 125 tests.

One 12.5 mL dispenser of DISCOVERY DAB CM contains a phosphate buffer and diaminobenzidine (DAB) and gentamicin sulfate.

One 12.5 mL dispenser of DISCOVERY H₂O₂ CM contains a phosphate buffer and hydrogen peroxide (H₂O₂).

One 12.5 mL dispenser of DISCOVERY Copper CM contains an acetate buffer and cupric sulfate and gentamicin sulfate.

One 12.5 mL dispenser of DISCOVERY Inhibitor CM contains hydrogen peroxide (H₂O₂).

Further dilution may result in loss of antigen staining. Differences in tissue processing and technical procedures in the laboratory may produce significant variability in results and require regular use of controls.

MATERIALS REQUIRED BUT NOT PROVIDED

Staining reagents, such as VENTANA detection kits and ancillary components, including negative and positive tissue control slides, are not provided.

STORAGE AND STABILITY

Upon receipt and when not in use, store at 2-8°C. Do not freeze.

To ensure proper reagent delivery and stability of the antibody, replace the dispenser cap after every use and immediately place the dispenser in the refrigerator in an upright position.

Every reagent is expiration dated. When properly stored, the reagent is stable to the date indicated on the label. Do not use reagent beyond the expiration date.

SPECIMEN PREPARATION

FFPE tissues are suitable for use with primary antibodies and/or probes when used with VENTANA detection kits and a DISCOVERY instrument. The recommended tissue fixative is 10% neutral buffered formalin.¹ Slides should be stained immediately, as antigenicity of cut tissue sections may diminish over time.

It is recommended that positive and negative controls should be run simultaneously with unknown specimens.

WARNINGS AND PRECAUTIONS

- For Research Use Only (RUO).
- For professional use only.
- Do not use beyond the specified number of tests.
- Positively charged slides may be susceptible to environmental stresses resulting in inappropriate staining. Ask your Roche representative for more information on how to use these types of slides.
- Materials of human or animal origin should be handled as biohazardous materials and disposed of with proper precautions. Consult local and/or state authorities for the recommended method of disposal. In the event of exposure, the health directives of the responsible authorities should be followed.^{2,3}
- Avoid contact of reagents with eyes and mucous membranes. If reagents come in contact with sensitive areas, wash with copious amounts of water.
- Avoid microbial contamination of reagents as it may cause incorrect results.

- For further information on the use of this device, refer to the instrument User Guide, and instructions for use of all necessary components located at navifyportal.roche.com.
- Consult local and/or state authorities with regard to the recommended method of disposal.
- Product safety labeling primarily follows EU GHS guidance. Safety data sheet available for professional user on request.
- To report suspected serious incidents related to this device, contact the local Roche representative and the competent authority of the Member State or Country in which the user is established.

This product contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

Table 1. Hazard information.

Hazard	Code	Statement
	H317	May cause an allergic skin reaction.
	H350	May cause cancer.
	H412	Harmful to aquatic life with long lasting effects.
	P201	Obtain special instructions before use.
	P261	Avoid breathing mist or vapours.
	P273	Avoid release to the environment.
	P280	Wear protective gloves/ protective clothing/ eye protection/ face protection/ hearing protection.
	P308 + P313	IF exposed or concerned: Get medical advice/ attention.
	P333 + P313	If skin irritation or rash occurs: Get medical advice/ attention.

INSTRUCTIONS FOR USE

Register the kit on your DISCOVERY series instrument, as described in the on-screen directions. Open the package, and take out the dispensers. Remove the cap from the nozzle of each dispenser, and place it on the nozzle cap holder on the rear of the dispenser. While holding the dispenser upright, remove the yellow shipping key by pulling the key tab to disengage it from each end. Do not cover the nozzle tip or depress the dispenser while removing key. Place the dispensers on the reagent tray, along with the appropriate accessory reagents.

The Discovery series of instruments and the ChromoMap DAB Kit form an integrated system. ALL KIT COMPONENTS MUST BE USED TOGETHER in order to obtain high-quality and consistent results. Omitting or changing any of the solutions may compromise the final outcome.

Bulk reagents should be prepared using quality reagent-grade water, not tap water. Carboys for storing bulk reagents should be rinsed thoroughly between fillings.

TROUBLESHOOTING

- If the positive control exhibits weaker staining than expected, other positive controls run during the same instrument run should be checked to determine if it is because of the primary antibody or one of the common secondary reagents.
- If the positive control is negative, it should be checked to ensure that the slide has the proper bar code label. If the slide is labeled properly, other positive controls run on the same instrument run should be checked to determine if it is because of the primary antibody or one of the common secondary reagents. Tissues may have been improperly collected, fixed, or deparaffinized. The proper procedure should be followed for collection, storage and fixation.
- If excessive background staining occurs, high levels of endogenous biotin may be present. Include a biotin blocking step in the staining protocol.
- If all of the paraffin has not been removed, there may be no staining. Repeat the deparaffinization procedure.
- If specific antibody staining is too intense, repeat the staining run and shorten the incubation time by 4 minute intervals until the desired stain intensity is achieved.

6. If tissue sections wash off the slide, slides should be checked to ensure that they are positively charged.
7. For corrective action, refer to the Instructions for Use section, the instrument User Guide or contact your local support representative.
8. If a reagent dispenser does not dispense fluid, check the priming chamber or meniscus for foreign materials or particulates, such as fibers or precipitates. If the dispenser is blocked, do not use the dispenser and contact your local support representative. Otherwise, re-prime the dispenser by aiming the dispenser over a waste container, removing the nozzle cap, and pressing down on the top of the dispenser.

REFERENCES

1. Carson FL, Cappellano C. Histotechnology; A Self-Instructional Text, 5th edition. American Society for Clinical Pathology Press; 2020, 2022.
2. Occupational Safety and Health Standards: Occupational exposure to hazardous chemicals in laboratories. (29 CFR Part 1910.1450). Fed. Register.
3. Directive 2000/54/EC of the European Parliament and Council of 24 June 2020 on the protection of workers from risks related to exposure to biological agents at work.

Symbols

Ventana uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see elabdoc.roche.com/symbols for more information).

GTIN Global Trade Item Number

REVISION HISTORY

Rev	Updates
E	Updates to Warnings and Precautions section.

INTELLECTUAL PROPERTY

VENTANA and DISCOVERY are trademarks of Roche. All other product names and trademarks are the property of their respective owners.

© 2024 Ventana Medical Systems, Inc.

CONTACT INFORMATION



Ventana Medical Systems, Inc.
 1910 E. Innovation Park Drive
 Tucson, AZ 85755
 USA
 +1 520 887 2155
 +1 800 227 2155 (USA)
www.roche.com