

Customer Address:

Customer Number : 2049990013
Customer Mat # : 06419232001
Roche Diagnostics Corporation
Accounts Payable
Po Box 50457
INDIANAPOLIS IN 46250-0417
UNITED STATES

Certificate of Analysis

Product Name : Uroplakin III (SP73) Rabbit Monoclonal Antibody
Product Number : 760-4533
Batch Number : V0002493
Source Batch : 0000104053
Storage Temperature : 2°C - 8°C
Expiry Date : 2023-08-31
Quality Release Date : 2021-02-18
Date of Manufacture : 2020-09-01

Test	Specification	Result
Appearance (Turbidity)	Free of Visible Foreign Material	Free of Visible Foreign Material
Appearance (Color)	Green Transparent	Green Transparent
Appearance (Form)	Liquid	Liquid
IHC Staining Intensity difference between sample and a reference must be less than or equal to 0.5.	Pass	Pass
IHC Staining Background difference between sample and a reference must be less than or equal to 0.5.	Pass	Pass
IHC - Tissue	Tissue Type	Urothelial Carcinoma
ELISA Titer	0.001 - 0.050 µg/ml	0.013 µg/ml
Bioburden Visible bacterial colonies must be less than 100 CFU/mL.	Pass	Pass
Source	Supernatant	Supernatant

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.



Product 760-4533
Batch V0002493



Patrick Strather, QC Inspector
Quality Assurance
Rocklin, California
US

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.

