



Certificate of Analysis

Product name

KIT COBAS 58/68/8800 EBV 192T IVD

Material No.

09040943190

Lot No.

N25472

Date of Manufacturing

2026-01-27

Expiry Date

2027-08

Parameter	Result	Unit
BGSD-ValidReps-Negativity	100	%
CAL HI-ValidReps-Acc-HitRate	100	%
CAL HI-ValidReps-Prec-SD	0.02	
CAL LO-ValidReps-Acc-HitRate	100	%
CAL LO-ValidReps-Prec-SD	0.04	
5X LOD-ValidReps-Sens-Pos	100	%
Visual Inspection	PASS	
Components	1	

This lot meets our specification

Disposition by

Tressa Dombroski

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2026-02-04

Page 1 of 1

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771