



## Certificate of Analysis

**Product name**

VENTANA ANTI-P120 CATENIN (98)

**Material No.**

05867088001

**Lot No.**

H35485

**Date of Manufacturing**

2021-12-13

**Expiry Date**

2023-12-12

**Parameter****Result****Unit**

Visual Inspection

Pass

Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Functional Staining

Pass

Ductal and Lobular Breast Carcinoma Tissues

The test batch and reference batch must score 2.0 or greater on both ductal and lobular breast carcinoma tissues for specific stain intensity on a 0 to 3 scale.

The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch.

The background intensity of the test batch, reference batch, and negative reagent control must be 0.5 or less on both tissues

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

**Disposition by**

Joshua Burgette

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2022-01-24

Page 1 of 1

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 800 227 2155  
<http://www.ventana.com>