



Certificate of Analysis

Product name

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

Material No.

05278368001

Lot No.

M25962

Date of Manufacturing

2025-01-07

Expiry Date

2026-06-17

Parameter**Result****Unit**Visual Inspection
Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Pass

Functional Staining

Pass

Breast Carcinoma Tissue
Gastric Carcinoma Tissue
Qualified 4 in 1 Cell Line ControlsThe specific staining of the test batch must be acceptable on all tissues and cell line cores.
The test batch background intensity must not exceed 0.5 pt. on any tissue or cell line core.
All tissues and cell line cores stained with the negative control reagent must be negative for specific staining.Bioburden
Visible bacterial colonies must be 100 or less CFU/mL

Pass

This lot meets our specification

Disposition by

Antonio Reyes

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-01-16

Page 1 of 1

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>