

CERTIFICATE OF ANALYSIS

Product: CONFIRM™ anti-p53 (DO7) Primary Antibody
 Ventana REF: 790-2912
 Roche Number: 05278074001
 Batch Number: E03348

Test	Acceptance Criteria	Result
Functional Staining on breast carcinoma or colon tissue	Reference slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be within 0.5	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2016-05-28

Manufacture date : 20-May-2014

Laura A. Jimenez 

22-MAY-2014

Quality Department

Date