

**CERTIFICATE OF ANALYSIS**


---

Product: Confirm™ anti-Progesterone Receptor (PR)(1E2) Rabbit Monoclonal Primary Antibody  
 Ventana REF: 790-2223  
 Roche Number: 05277990001  
 Batch Number: E04944

Parameter	Result	
Functional Staining on breast carcinoma tissue and reference tissue slide	Nuclear staining must be present in 51% or more of the tumor cells	PASS
	Background staining reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be less than or equal to 0.5	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2016-06-28

Manufacture date: 09-JUL-2014

Lucia Romero  09-JUL-2014  
 \_\_\_\_\_  
**Quality Department** **Date**