



## Certificate of Analysis

**Product name**

VENTANA ANTI-P63 (4A4)

**Material No.**

05867061001

**Lot No.**

G23734

**Date of Manufacturing**

2020-10-21

**Expiry Date**

2022-10-20

**Parameter****Result****Unit**

Final Acceptance Testing

Pass

Anti-p63 (4A4) Mouse Monoclonal Primary Antibody stains the nuclei of human prostatic basal cells in normal regions of prostate carcinoma and squamous epithelial cells in the tonsil.

The specific staining of the Anti-p63 (4A4) Mouse Monoclonal Primary Antibody test lot must score 3.0 or greater in normal cells of the prostate carcinoma tissue.

The neoplastic cells will be used as an intra-tissue control, and specific stain intensity must score less than or equal to 0.5.

Specific staining in the tonsil must score 3.0 or greater on squamous epithelial cells.

The test lot specific stain intensity must also be within 0.5 pt. of the reference lot for a passing result.

The intensity of the negative reagent control slide must be 0.5 or less.

**Issue date** 2020-11-06

Page 1 of 2

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 520 887 2155  
<http://www.ventana.com>



## Certificate of Analysis

**Product name**

VENTANA ANTI-P63 (4A4)

**Material No.**

05867061001

**Lot No.**

G23734

**Date of Manufacturing**

2020-10-21

**Expiry Date**

2022-10-20

**Parameter****Result****Unit**

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

This lot meets our specification

**Disposition by**

Antonio Reyes

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2020-11-06

Page 2 of 2

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 520 887 2155  
<http://www.ventana.com>