

CERTIFICATE OF ANALYSIS

Product Name: Anti-Pan Keratin (AE1/AE3/PCK26) Primary Antibody
 Ventana REF : 760-2595
 Roche Number: 05267145001
 Batch Number: D03295

Test	Acceptance Criteria	Results
Functional Staining on skin, liver, breast cancer, prostate cancer, and colon cancer. May use multi tissue block if available	Skin slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Liver slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Breast cancer slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Prostate cancer slide must score 3.5+ or greater on a scale of 0-4+	PASS
	Colon cancer slide must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be 0.5+ or less on a scale of 0-4+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2015-05-28

Manufacture date : 30-MAY-2013

Elizabeth Garcia

Elizabeth Garcia 31-May-2013

Quality Department

Date