

Certificate of Analysis

Product name

KIT COBAS 58/68/8800 HIV 192T IVD

Material No.

09040803190

 Lot No.
 Date of Manufacturing
 Expiry Date

 M12542
 2024-10-02
 2026-06

Parameter	Result	Unit
ERCB-Accuracy	100	%
ERCB-Prec-SD	0.09	
ERCD-Accuracy	100	%
ERCD-Prec-SD	0.07	
5XLLOQ-Sens Reps-Pos	100	%
NHP-Negativity-TND	100	%
Components	1	

This lot meets our specification

Disposition by

Aditi Shah

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2024-10-15 Page 1 of 1

Roche Molecular Systems, Inc. Quality Management 1080 U.S. Highway 202 South Branchburg, NJ 08876-3771