



a member of the Roche group

CERTIFICATE OF ANALYSIS

Product Name: VENTANA anti-E-Cadherin (36) Mouse Monoclonal Primary Antibody

Ventana REF: 790-4497

Roche Number: 05905290001


Batch Number: F09582

Test	Acceptance Criteria	Result
Functional Staining on ductal breast and lung carcinoma	The test batch must score 2.5+ or greater on ductal breast and lung carcinoma tissue	PASS
	The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result	PASS
	The background intensity must be 0.5 or less	PASS
	The intensity of the negative reagent control slide must be 0.5 or less	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiration date: 2017/12/28

Manufacture date: 14/Dec/2015


Jeremy Thomas
22 . JAN . 2016

Quality Department
Date