

Certificate of Analysis

Product Name: Glial Fibrillary Acidic Protein (EP672Y) Rabbit Monoclonal Antibody
Catalog Number(s): 760-4345
Roche Part Number: 05269784001
Lot Number: V0000646
Date of Manufacture: 2016-05-12
Expiration Date: 2019-04-30

Test	Acceptance Criteria	Results
Functional testing on the following tissue(s): Brain	Intensity difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
	Background difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Appearance	Transparent liquid and free of visible foreign material.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Bioburden	Visible bacterial colonies must be less than 100CFU/mL.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
ELISA	1.0-10.0 µg/mL	1.19 µg/mL
Source	N/A	Supernatant

Additional information: N/A



Authorized Signature:  Date: 7/27/17
Quality Assurance Representative