



Certificate of Analysis

Product name

KIT C58/68/88 HIV-1/HIV-2 QUAL 192T IVD

Material No.

09040528190

Lot No.

H32139

Date of Manufacturing

2022-01-07

Expiry Date

2023-04

Parameter	Result	Unit
HIV1O-2.5X-Sens/Repeat Reps-R	100	%
HIV2-2.5X-Sens/Repeat Reps-R	100	%
HIV1O-0.5X-SensReps-R	76	%
HIV2-0.5X-SensReps-R	76	%
NHP-Negativity-NR	100	%
MBC 0.5X-HIV1M-Ch1-SensReps-R	90	%
HIV1M-2.5X-Sens/Repeat Reps-R	100	%
Components	1	

This lot meets our specification

Disposition by

Craig Cawley

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2022-02-17

Page 1 of 1

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771