



## Certificate of Analysis

**Product name**

KIT COBAS 4800 CMV 120T CE-IVD

**Material No.**

07865970190

**Lot No.**

E18902

**Date of Manufacturing**

2018-10-18

**Expiry Date**

2019-10

Parameter	Result	Unit
Components	1	
5XLOD-Sens Reps-Pos	100	%
C48CMVKIT-120T	PASS	

The Visual Inspection should be as follows:

10 X ADAPTER c4800 CMV MMX R2, 0.5 mL  
Black microtube and cap with a gold insert, fitted firmly into a labeled adapter.

10 X ADAPTER c4800 MMX R1, 1.75 mL  
Natural microtube and cap with a teal insert, fitted firmly into a labeled adapter.

10 X ADAPTER c4800 DNA QS, 1.75 mL  
Natural microtube and cap with a purple insert, fitted firmly into a labeled adapter.

Calibration Coefficient Verification  
Kit Calibration Coefficient values encoded in the Assay

**Issue date** 2018-10-24

Page 1 of 2

Roche Molecular Systems, Inc.  
Quality Management  
1080 U.S. Highway 202 South  
Branchburg, NJ 08876-3771



## Certificate of Analysis

**Product name**

KIT COBAS 4800 CMV 120T CE-IVD

**Material No.**

07865970190

**Lot No.**

E18902

**Date of Manufacturing**

2018-10-18

**Expiry Date**

2019-10

**Parameter****Result****Unit**

Value Card must match the values recorded in SAP.

CALHI1-Accuracy	100	%
CALHI1-Prec-SD	0.02	
STDLO-Accuracy	100	%
STDLO-Prec-SD	0.04	
NHP-Negativity-NR	100	%

This lot meets our specification

**Disposition by**

Mujib Walli

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2018-10-24

Page 2 of 2

Roche Molecular Systems, Inc.  
Quality Management  
1080 U.S. Highway 202 South  
Branchburg, NJ 08876-3771