



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

F27098

Date of Manufacturing

2019-12-04

Expiry Date

2021-11-25

Parameter**Result****Unit**

Final Acceptance Testing

Pass

Anti-p53 (Bp53-11) Primary Antibody produces nuclear punctate staining of reactive carcinoma cells.

The test batch of anti-p53 (Bp53-11) Antibody must score 3.5+ or greater on colon or ovarian carcinoma tissue with a background intensity of 0.5+ or less.

The specific stain intensity must also be within 0.5 pt.

The intensity of the negative reagent control slide must be 0.5 or less.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

This lot meets our specification

Issue date 2019-12-17

Page 1 of 2



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

F27098

Date of Manufacturing

2019-12-04

Expiry Date

2021-11-25

Disposition by

Mia Cazares

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2019-12-17

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>