

Elecsys ApoE4p RUO

REF			SYSTEM
09697888430	09697888520	100	cobas e 402 cobas e 801

English

For Research Use Only (RUO). Not for use in diagnostic procedures.

System-Information

Short name	ACN (application code number)
ApoE4PQv	17055

Please ensure installation of the lot-specific efiles on the **cobas** link prior to running the assay.

Please note

The measured value of ApoE4 in a given sample, determined with assays from different manufacturers, can vary due to differences in assay methods and reagents.

Assay description

The Elecsys ApoE4p RUO Assay is a qualitative immunoassay for the in vitro determination of Apolipoprotein E4 (ApoE4) in human plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.

Test principle

The Elecsys ApoE4p RUO assay is a qualitative, 2-incubation step assay using the sandwich assay principle.

Total duration of assay: 18 minutes.

- 1st incubation: 30 µL of the automatically diluted sample (using Diluent Universal), a biotinylated monoclonal ApoE-specific antibody (R1) and a monoclonal ApoE4-specific antibody (R2) labeled with a ruthenium complex^{a)} react to form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell II M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined automatically by the software by comparing the electrochemiluminescence signal obtained from the reaction product of the sample with the signal of the cutoff value previously obtained by calibration.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The **cobas e** pack (M, R1, R2) is labeled as ApoE4PQv.

- M Streptavidin-coated microparticles, 1 bottle, 5.8 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-ApoE-Ab-biotin, 1 bottle, 9 mL:
Fab-fragment of a biotinylated monoclonal anti-ApoE antibody (mouse), 3.61 mg/L; phosphate buffer 100 mmol/L, pH 7.4; preservative
- R2 Anti-ApoE4-Ab-Ru(bpy)₃²⁺, 1 bottle, 9 mL:
Monoclonal anti-ApoE4 antibody (mouse) labeled with ruthenium complex, 1.1 mg/L; phosphate buffer 100 mmol/L, pH 7.4; preservative

Warnings and precautions

Do not run RUO assays in random access mode with IVD measurements. Instead, perform RUO assays in batch mode and follow the washing procedure to avoid any potential impact on IVD results.

Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Warning

H317 May cause an allergic skin reaction.

H412 Harmful to aquatic life with long lasting effects.

Prevention:

P261 Avoid breathing mist or vapours.

P273 Avoid release to the environment.

P280 Wear protective gloves.

Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Hazardous components:

- 2-methyl-2H-isothiazol-3-one hydrochloride
- Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

All human material should be considered potentially infectious.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Handling

cobas e pack:

The reagents in the kit (M, R1 and R2) are ready-for-use and are supplied in bottles compatible with the system.

All information required for correct operation is read in from the respective reagent labels or can be found in the package insert and on the associated Value Sheet.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the **cobas e** pack **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date (see Value Sheet)
on the analyzers	112 days

Elecsys ApoE4p RUO



Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

EDTA plasma.

Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 15-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Close the cap when reagents are not in use.

Materials provided

See "Reagents – working solutions" section for reagents.

cobas e pack (M, R1, R2) ApoE4pQv.

Barcode labels for the Diluent Universal, 36 mL sample diluent

[REF] 07299001190.

Materials required (but not provided)

- [REF] 09697900430, CalSet ApoE4p RUO, for 2 x 1.0 mL
- [REF] 09697934430, PreciControl ApoE4p RUO, for 2 x 1.0 mL
- [REF] 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- [REF] 03142949122, ControlSet Vials, 2 x 56 empty snap-cap bottles
- [REF] 07299001190, Diluent Universal, 36 mL sample diluent
- Distilled or deionized water
- General laboratory equipment
- **cobas e** analyzer

Additional materials for **cobas e** 402 and **cobas e** 801 analyzers:

- [REF] 06908799190, ProCell II M, 2 x 2 L system solution
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 07485409001, Reservoir Cup, 8 cups to supply ProCell II M and CleanCell M
- [REF] 06908853190, PreClean II M, 2 x 2 L wash solution
- [REF] 05694302001, Assay Tip/Assay Cup tray, 6 magazines x 6 magazine stacks x 105 assay tips and 105 assay cups, 3 wasteliners
- [REF] 07485425001, Liquid Flow Cleaning Cup, 2 adaptor cups to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning Detection Unit
- [REF] 07485433001, PreWash Liquid Flow Cleaning Cup, 1 adaptor cup to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning PreWash Unit
- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution
- [REF] 11298500160, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution (for USA)

Test procedure

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use.

Place the cooled (stored at 2-8 °C) **cobas e** pack on the reagent manager. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the **cobas e** pack.

Read the test-specific parameters via the reagent label.

Automatic pre-dilution of sample

The Elecsys ApoE4p RUO assay automatically performs a pre-dilution of samples using Diluent Universal as part of the 1st incubation step (section "Test principle").

In order for the analyzer to recognize the Diluent Universal (configured within the assay efiles), the specific barcode labels provided (section "Materials provided") must be applied to the Diluent Universal prior to being loaded onto the analyzer.

Calibration

Traceability: No WHO standard is available for ApoE4p.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the **cobas e** pack was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- every 12 weeks when using the same reagent lot
- every 28 days when using the same **cobas e** pack on the analyzer
- as required: e.g. quality control findings outside the defined limits

Quality control

Use PreciControl ApoE4p RUO or other suitable controls for routine quality control procedures.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Note: The controls are not barcode-labeled and therefore have to be run like external controls. All values and ranges have to be entered manually. Please refer to the section "QC" in the operator's manual or to the online help of the instrument software.

Non-barcode labeled controls: Only one target value and range for each control level can be entered in the analyzer. The reagent lot-specific target values have to be re-entered each time a specific reagent lot with different control target values and ranges is used. Two reagent lots with different control target values and ranges cannot be used in parallel in the same run.

The exact lot-specific target values and ranges are printed on the enclosed (or electronically available) value sheet in the reagent kit or PreciControl kit. Please make sure that the correct values are used.

Calculation

The analyzer automatically calculates the cutoff based on the measurement of Cal1 ApoE4p and Cal2 ApoE4p. The result of a sample is given as reactive or non-reactive as well as in the form of a cutoff index (COI; signal sample/cutoff).

For further information, please refer to the appropriate user guide or operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see navifyportal.roche.com for definition of symbols used):

CONTENT Contents of kit

Elecsys ApoE4p RUO



SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume for reconstitution

COBAS, ELECSYS, NAVIFY and PRECICONTROL are trademarks of Roche.
All other product names and trademarks are the property of their respective owners.
Additions, deletions or changes are indicated by a change bar in the margin.
© 2025, Roche Diagnostics

 Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany
www.roche.com
 +800 5505 6606

