



Certificate of Analysis

Product name
Anti-p63 (4A4) Mouse Monoclonal Primary Antibody

Material No.
05867061001

Lot No.	Date of Manufacturing	Expiry Date
E10888	14-May-2018	2020-05-13

Parameter	Result	Unit
------------------	---------------	-------------

Visual Inspection	Pass	
-------------------	------	--

Reagent must be free of foreign material and substantially free of particulate matter

Reagent must be clear and free of turbidity

Final Acceptance Testing	Pass	
--------------------------	------	--

Prostate carcinoma and tonsil

The test batch must score 3 or greater on prostate carcinoma and tonsil tissue

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result

The background intensity must be within 0.5 of the negative control slide

The intensity of the negative reagent control slide must be 0.5 or less

Bioburden	Pass	
-----------	------	--

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

WHITNEY CROFT WC

Issue date

JUN 2018