



Certificate of Analysis

Product name

BRAV V600E (VE1) MM PAB-US Export

Material No.

08033706001

Lot No.

E03407

Date of Manufacturing

2018-02-07

Expiry Date

2020-02-06

Parameter**Result****Unit**

Final Acceptance Testing

Pass

Positive Colon Carcinoma

The test batch of VENTANA anti-BRAF V600E (VE1) Mouse Monoclonal Primary Antibody must score 1.0 point or greater on colon CRC tissue with acceptable background.

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result.

The background intensity must be equal to or less than 0.5 point on a 0-3 scale.

The intensity of the negative reagent control slide must be 0.5 point or less on a 0-3 scale.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Issue date 2018-02-19

Page 1 of 2



Certificate of Analysis

Product name

BRAV V600E (VE1) MM PAB-US Export

Material No.

08033706001

Lot No.

E03407

Date of Manufacturing

2018-02-07

Expiry Date

2020-02-06

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

This lot meets our specification

Disposition by

Whitney Croft

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2018-02-19

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>