



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

K29345

Date of Manufacturing

2024-01-23

Expiry Date

2026-01-11

Parameter**Result****Unit**

Functional Staining

Pass

Colon carcinoma or Ovarian carcinoma

The anti-p53 (Bp53-11) Primary Antibody produces nuclear punctate staining of reactive carcinoma cells.

For specific stain intensity on a 0 to 4 scale, the test batch and reference batch must score 3.5 pt or greater in Colon carcinoma or Ovarian carcinoma.

The background intensity must be 0.5 pt or less for the test batch and reference batch.

The run is invalid if the reference batch does not exhibit acceptable stain intensity.

The background intensity of the negative reagent control slide must be 0.5 pt or less.

Visual Inspection

Pass

Reagent must be clear of foreign material and particulate matter, and free of turbidity

Bioburden

Pass

Issue date 2024-02-06

Page 1 of 2



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

K29345

Date of Manufacturing

2024-01-23

Expiry Date

2026-01-11

Parameter**Result****Unit**

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Mia Cazares

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2024-02-06

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>