

CERTIFICATE OF ANALYSIS

Product Name: Protease 1
Ventana REF : 760-2018
Roche Number: 05266688001
Batch Number: F05109

Test	Acceptance Criteria	Result
Spectrophotometer testing	Mean Absorbance must score in the range of 0.50 and 1.30	PASS
Visual Inspection	Reagent should be clear and colorless, but it is acceptable for the reagent to contain a white flocculent material of a non-biological origin.	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL.	PASS

This lot met the specifications.

Expiry Date: 2017/05/28

Manufacture Date: 30-JUL-2015

Quality Department: Laughlin Hoskinson Laughlin Hoskinson Date: 26-Aug-2015

Rev S