



a member of the Roche group

CERTIFICATE OF ANALYSIS

Product Name: Anti-p53 (BP53-11) Primary Antibody
Ventana **REF**: 760-2542
Roche Number: 05267102001
Batch Number: F08699

Test	Acceptance Criteria	Result
Functional Staining on colon or ovarian carcinoma tissue	Test batch slide reading must score 3.5+ or greater (scale 0-4+)	PASS
	Background reading must score 0.5+ or less	PASS
	Intensity difference between sample and a reference must be less than or equal to 0.5+	PASS
	Intensity of the negative reagent control slide must be 0.5 or less	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry Date: 2017-11-28

Manufacture Date: 30-NOV-2015

Guadalupe Villarreal

Guadalupe Villarreal 18 Dec 2015

Quality Department

Date