



Certificate of Analysis

Product name

CONFIRM ANTI-PR (1E2)

Material No.

05277990001

Lot No.

N00777

Date of Manufacturing

2025-03-06

Expiry Date

2027-02-19

Parameter**Result****Unit**

Functional Staining

Pass

Breast Carcinoma Tissue & Normal Breast Tissue

Nuclear staining will be interpreted on Breast Carcinoma that is positive for percent tumor cell staining according to the following approximate percent tumor cell staining criteria:

0-50% cell staining (negative)
51-100% cell staining (positive).

Nuclear staining of test and reference breast carcinoma slides must have a specific stain intensity of 3.5 or greater and a background of 0.5 or less on a 0 to 4 scale.

The test batch on normal breast slide must exhibit a heterogeneous staining pattern of the luminal cells, with a mixture of a variable number of cells exhibiting weak, moderate, and intense immunoreactivity.

The background intensity of the negative reagent control slide must be 0.5 or less.

Visual Inspection

Pass

Issue date 2025-03-18

Page 1 of 2



Certificate of Analysis

Product name

CONFIRM ANTI-PR (1E2)

Material No.

05277990001

Lot No.

N00777

Date of Manufacturing

2025-03-06

Expiry Date

2027-02-19

Parameter**Result****Unit**

Reagent must be clear of foreign material and particulate matter, and free of turbidity

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Nic Hau

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-03-18

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>