



Certificate of Analysis

Product name

KIT COBAS 4800 HCV 120T CE-IVD

Material No.

06979602190

Lot No.

M16484

Date of Manufacturing

2025-02-11

Expiry Date

2026-07

Parameter	Result	Unit
Components	1	
5X LOD-Sens Repts-Pos	100	%
C4800-HCVKIT-120T	PASS	

The visual inspection should be as follows:

10 X ADAPTER c4800 HCV MMX R2, 0.5 mL

Black microtube and cap with a gold insert, containing a clear, faint blue to faint purple liquid, fitted firmly into a labeled adapter.

10 X ADAPTER c4800 MMX R1, 1.75 mL

Natural microtube and cap with a teal insert, containing a clear, colorless liquid, fitted firmly into a labeled adapter.

10 X ADAPTER c4800 RNA QS, 1.75 mL

Natural microtube and cap with a purple insert, containing a clear colorless liquid, fitted firmly into a labeled adapter.

Calibration Coefficient Verification

Kit Calibration Coefficient values encoded in the Assay Value Card must match the values recorded in SAP.

Issue date 2025-03-13

Page 1 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771



Certificate of Analysis

Product name

KIT COBAS 4800 HCV 120T CE-IVD

Material No.

06979602190

Lot No.

M16484

Date of Manufacturing

2025-02-11

Expiry Date

2026-07

Parameter	Result	Unit
HC-Accuracy	100	%
HC-Prec-SD	0.03	
LC-Accuracy	100	%
LC-Prec-SD	0.07	
NHP-Negativity-TND	100	%

This lot meets our specification

Disposition by

Theresa Lapinski

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-03-13

Page 2 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771