

**CERTIFICATE OF ANALYSIS**


---

Product Name: VENTANA anti-HER2/neu (4B5) Rabbit Monoclonal Primary Antibody  
 Ventana [REF]: 790-4493  
 Roche Number: 05999570001  
 Batch Number: E03342

Test	Acceptance Criteria	Result
Functional Staining on breast and gastric carcinoma tissues	Test batch must score 3+ or greater on both gastric and breast carcinoma tissue	PASS
	Background intensity must be within 0.5 of the negative control slide	PASS
	Specific stain intensity of test batch must be within 0.5 of the reference batch	PASS
	Intensity of the negative reagent control slide must be 0.5 or less	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2015-10-28

Manufacture date : 13-MAY-2014



Patricia Gomez

14-MAY-2014

**Quality Department**

**Date**