



Certificate of Analysis

Product Name: CD1a (EP3622) Rabbit Monoclonal Antibody
Catalog Number(s): 760-4525
Roche Part Number: 06419160001
Lot Number: V0000821
Date of Manufacture: 2017-07-12
Expiration Date: 2020-06-30

Test	Acceptance Criteria	Results
Functional testing on the following tissue(s): Thymus	Intensity difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
	Background difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Appearance	Transparent liquid and free of visible foreign material.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Bioburden	Visible bacterial colonies must be less than 100CFU/mL.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
ELISA	0.1-5.0 µg/mL	2.17 µg/mL
Source	N/A	Supernatant

Additional information: N/A



Authorized Signature: Date: 10/12/17
Quality Assurance Representative

Issued: 6/3/16 Rev 1.0

Document #QF0901-008-02 [CM Legacy #]

Manufactured by: Cell Marque Corporation
6600 Sierra College Boulevard • Rocklin, California 95677 • USA • Telephone (916) 746-8900 • Fax (916) 746-8976