



Certificate of Analysis

Product name

OptiView Amplification Kit (250)

Material No.

06718663001

Lot No.

E04580

Date of Manufacturing

2018-03-07

Expiry Date

2019-08-21

Parameter**Result****Unit**

Final Acceptance Testing

Acceptable

Tonsil tissue and reference tissue slide

Antibody staining with Amplification Kit must be assigned a staining intensity value of +1.0 or greater than those slides stained with OptiView DAB Detection Kit.

Background will only be evaluated on the #No antibody control slides and must be equivalent between OptiView Amplification Kit and OptiView DAB Detection Kit.

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

This lot meets our specification

Issue date 2018-03-17

Page 1 of 2



Certificate of Analysis

Product name

OptiView Amplification Kit (250)

Material No.

06718663001

Lot No.

E04580

Date of Manufacturing

2018-03-07

Expiry Date

2019-08-21

Disposition by

Ken Eblen

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2018-03-17

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>