



Certificate of Analysis

Product name

KIT COBAS CFDNA SAMP PREP 24T

Material No.

07247737190

Lot No.

N00153

Date of Manufacturing

2025-04-08

Expiry Date

2026-08

Parameter**Result****Unit**

Components

1

CFDNA-SAMPPREP

PASS

The visual inspection should be as follows:

8 x DNA PBB (DNA Paraffin Binding Buffer, 10 mL)
Natural bottle and screw cap containing a clear,
colorless liquid

1 x WB I (Wash Buffer I, 25 mL)
Natural bottle and screw cap containing a clear,
colorless liquid

1 x WB II (Wash Buffer II, 12.5 mL)
Natural bottle and screw cap containing a clear,
colorless liquid

1 x DNA EB (DNA Elution Buffer, 6 mL)
Natural bottle and screw cap containing a clear,
colorless liquid

2 x PK (Proteinase K, 100 mg)
Brown glass bottle containing a white, clumpy powder

5 x HPEA FT (High Pure Extension Assembly Unit, 5

Issue date 2025-04-21

Page 1 of 3

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771



Certificate of Analysis

Product name

KIT COBAS CFDNA SAMP PREP 24T

Material No.

07247737190

Lot No.

N00153

Date of Manufacturing

2025-04-08

Expiry Date

2026-08

Parameter**Result****Unit**

pcs)
Bag containing 5 assembled filter tube with cap and
extender in 50 mL centrifugation tubes

3 x CT (Collection Tubes, 25 pcs)
Bag containing 25 polypropylene tubes

K562 DNA-MMX1-Ch5-Ct

4

#InRng

K562 DNA-MMX2-Ch5-Ct

4

#InRng

K562 DNA-MMX3-Ch5-Ct

4

#InRng

PNC-MMX1-Ch5-Ct

4

#InRng

PNC-MMX2-Ch5-Ct

4

#InRng

PNC-MMX3-Ch5-Ct

4

#InRng

This lot meets our specification

Issue date 2025-04-21

Page 2 of 3

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771



Certificate of Analysis

Product name

KIT COBAS CFDNA SAMP PREP 24T

Material No.

07247737190

Lot No.

N00153

Date of Manufacturing

2025-04-08

Expiry Date

2026-08

Disposition by

Christine Penyak

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-04-21

Page 3 of 3

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771