

CERTIFICATE OF ANALYSIS

Product Name: CONFIRM anti-Topoisomerase II α (JS5B4) Rabbit Monoclonal Primary Antibody
 Ventana REF: 790-4371
 Roche Number: 05479339001
 Batch Number: E07638

Test	Acceptance Criteria	Result
Functional Staining on breast carcinoma tissue and reference tissue slide	Reference slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be less than or equal to 0.5+	PASS
	Test slide must not show a decrease in positive cell staining of more than 10% as compared to the reference slide.	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2016-09-28

Manufacture date : 25-SEP-2014

Elizabeth Garcia



29-Sep-2014

Quality Department

Date