

cobas e 411 analyzer

Addendum 1.0 to Safety Manual V1.2

Software version 03-01



Publication information

Publication version	Software version	Revision date	Change description
1.0	03-01	2019-09	First version. Safety message about sharp edges has been added to the Safety Manual.

☰ Revision history

Edition notice

This publication is for users of the **cobas e 411** analyzer.

Every effort has been made to ensure that all the information contained in this publication is correct at the time of publishing. However, the manufacturer of this product may need to update the publication information as output of product surveillance activities, leading to a new version of this publication.

Copyright

© 2019, Roche Diagnostics GmbH.
All rights reserved.

Trademarks

The following trademarks are acknowledged:

COBAS, COBAS C, COBAS E, and ELECSYS are trademarks of Roche.

All other trademarks are the property of their respective owners.


Contact addresses

Inside the European Union and EFTA member states



Manufacturer of
cobas e 411
instrument

Hitachi High-Technologies
Corporation
1-24-14 Nishi-Shimbashi
Minato-ku Tokyo 105-8717
Japan



Authorized
representative

Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim
Germany

Outside the European Union and EFTA member states

Manufactured by:

Hitachi High-Technologies
Corporation

Manufactured for:

Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim
Germany

Warning messages

Biohazardous materials

Sharps, rough edges, and/or moving parts

Personal injury and infection due to sharps, rough edges, and/or moving parts

Good Laboratory Practice can reduce the risk of injury. Be aware of your laboratory environment, well-prepared, and follow the instructions for use.

Some areas of the instrument may have sharps, rough edges, and/or moving parts.

- ▶ Wear personal protective equipment to minimize the risk of injury from bodily contact with such parts, especially in less accessible areas, or while cleaning the instrument.
- ▶ Your personal protective equipment should be appropriate to the degree and type of potential hazard, e.g. suitable lab gloves, eye protection, lab coat, and footwear.