



Certificate of Analysis

Product name

KIT COBAS 58/68/8800 BKV 192T IVD

Material No.

09040960190

Lot No.

N05657

Date of Manufacturing

2025-05-02

Expiry Date

2026-12

Parameter	Result	Unit
BGSD-ValidReps-Negativity	100	%
CAL HI-ValidReps-Acc-HitRate	100	%
CAL HI-ValidReps-Prec-SD	0.03	
CAL LO-ValidReps-Acc-HitRate	100	%
CAL LO-ValidReps-Prec-SD	0.05	
5X LOD-ValidReps-Sens-Pos	100	%
Components	1	

This lot meets our specification

Disposition by

Christine Penyak

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-05-14

Page 1 of 1

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771