



Certificate of Analysis

Product name

CONFIRM anti-Topoisomerase II (JS5B4)

Material No.

05479339001

Lot No.

N09792

Date of Manufacturing

2025-05-28

Expiry Date

2027-05-21

Parameter**Result****Unit**

Visual Inspection
Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Pass

Functional Staining

Pass

Breast CA

For specific stain intensity on a 0 to 4 scale, the test batch and reference batch must score 3.5 pt or greater in breast cancer tissue.

The background intensity must be 0.5 pt or less for the test batch and reference batch.

The background intensity of the negative reagent control must be 0.5 pt or less.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Joshua Burgette

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-06-23

Page 1 of 1