



# Certificate of Analysis

**Product name**

VENTANA ANTI-CYCLIN D1 (SP4-R)

**Material No.**

05862949001

**Lot No.**

K05667

**Date of Manufacturing**

2023-04-04

**Expiry Date**

2025-04-03

**Parameter****Result****Unit**

Visual Inspection

Pass

Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Functional Staining

Pass

Normal Tonsil and Breast Carcinoma Tissues

The test batch and reference batch must be 2 or greater for specific stain intensity on normal tonsil and breast carcinoma tissues on a 0 to 4 scale.

The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch.

The background intensity must be 0.5 or less on both tissues for the test batch and reference batch.

The background intensity of the negative reagent control on both tissues must be 0.5 or less.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

**Issue date** 2023-07-13

Page 1 of 2

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 800 227 2155  
<http://www.ventana.com>



# Certificate of Analysis

**Product name**

VENTANA ANTI-CYCLIN D1 (SP4-R)

**Material No.**

05862949001

**Lot No.**

K05667

**Date of Manufacturing**

2023-04-04

**Expiry Date**

2025-04-03

**Disposition by**

Joshua Burgette

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2023-07-13

Page 2 of 2

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 800 227 2155  
<http://www.ventana.com>