



Certificate of Analysis

Product name

KIT COBAS 58/68/8800 HCV 192T IVD

Material No.

09040765190

Lot No.

K04951

Date of Manufacturing

2023-07-12

Expiry Date

2025-01

Parameter	Result	Unit
5XLLOQ-Sens Reps-Pos	100	%
NHP-Negativity-TND	100	%
LC-Accuracy	100	%
LC-Prec-SD	0.05	
HC-Accuracy	100	%
HC-Prec-SD	0.03	
Components	1	

This lot meets our specification

Disposition by

Maria Simon

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2023-07-26

Page 1 of 1

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771