



Certificate of Analysis

Product Name: anti-PAX8 (MRQ-50) Mouse Monoclonal Primary Antibody
Catalog Number(s): 760-4618
Roche Part Number: 06523927001
Lot Number: V0001064
Date of Manufacture: 2017-06-20
Expiration Date: 2020-05-31

Test	Acceptance Criteria	Results
Functional testing on the following tissue(s): Ovarian carcinoma	Intensity difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
	Background difference between sample and a reference must be less than or equal to 0.5	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Appearance	Transparent liquid and free of visible foreign material.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
Bioburden	Visible bacterial colonies must be less than 100CFU/mL.	<input checked="" type="checkbox"/> PASS <input type="checkbox"/> FAIL
ELISA	5-15 µg/mL	5.60 µg/mL
Source	N/A	Supernatant

Additional information: N/A



Authorized Signature:  Date: 01/23/2018
Quality Assurance Representative

Issued: 6/3/16 Rev 1.0

Document #QF0901-008-02 [CM Legacy #]

Manufactured by: Cell Marque Corporation
6600 Sierra College Boulevard • Rocklin, California 95677 • USA • Telephone (916) 746-8900 • Fax (916) 746-8976