



Certificate of Analysis

Product name

CONFIRM anti-Topoisomerase II (JS5B4)

Material No.

05479339001

Lot No.

H25779

Date of Manufacturing

2021-10-08

Expiry Date

2023-09-09

Parameter**Result****Unit**

Final Acceptance Testing

Pass

Breast Cancer

The test batch must score 3.5 or greater on breast cancer tissue

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result

The background intensity must be within 0.5 of the negative control slide

The intensity of the negative reagent control slide must be 0.5 or less

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and

Issue date 2021-11-02

Page 1 of 2



Certificate of Analysis

Product name

CONFIRM anti-Topoisomerase II (JS5B4)

Material No.

05479339001

Lot No.

H25779

Date of Manufacturing

2021-10-08

Expiry Date

2023-09-09

Parameter**Result****Unit**

substantially free of intrinsic particulate matter.

This lot meets our specification

Disposition by

Joshua Burgette

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2021-11-02

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>