



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

Y15702

Date of Manufacturing

2017-08-09

Expiry Date

2019-07-26

Parameter**Result****Unit**

Functional Staining

Pass

Anti-p53 (Bp53-11) Primary Antibody produces nuclear punctate staining of reactive carcinoma cells.

The test batch of anti-p53 (Bp53-11) Antibody must score 3.5+ or greater on colon or ovarian carcinoma tissue with a background intensity of 0.5+ or less.

The specific stain intensity must also be within 0.5 pt.

The intensity of the negative reagent control slide must be 0.5 or less.

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter

Bioburden

Pass

Issue date 2017-09-01

Page 1 of 2



Certificate of Analysis

Product name

ANTI-P53 (Bp53-11) PaB

Material No.

05267102001

Lot No.

Y15702

Date of Manufacturing

2017-08-09

Expiry Date

2019-07-26

Parameter**Result****Unit**

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Whitney Croft

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2017-09-01

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>