



Certificate of Analysis

Product name

VENTANA anti-E-Cadherin (36) Mouse Monoclonal Primary Antibody

Material No.

05905290001

Lot No.

G09146

Date of Manufacturing

28-Nov-2016

Expiry Date

2018-11-28

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be free of foreign material and substantially free of particulate matter

Reagent must be clear and free of turbidity

Functional Staining

Pass

Ductal breast and lung carcinoma

The test batch must score 2.5+ or greater on ductal breast and lung carcinoma tissue

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result

The background intensity must be 0.5 or less

The intensity of the negative reagent control slide must be 0.5 or less

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Jeremy Thomas

Issue date

DEC/2016