

Customer Address:

Customer Number : 2049990013
Customer Mat # : 05269784001
Roche Diagnostics Corporation
Accounts Payable
Po Box 50457
INDIANAPOLIS IN 46250-0417
UNITED STATES

Certificate of Analysis


Product Name : Glial Fibrillary Acidic Protein (EP672Y) Rabbit Monoclonal Antibody
Product Number : 760-4345
Batch Number : V0004720
Source Batch : 0000355702
Storage Temperature : 2 - 8 °C
Expiry Date : 2027-05-31
Quality Release Date : 2024-08-01
Date of Manufacture : 2024-06-26

Test	Specification	Result
Appearance (Turbidity)	Free of Visible Foreign Material	Free of Visible Foreign Material
Appearance (Color)	Green Transparent	Green Transparent
Appearance (Form)	Liquid	Liquid
IHC Staining	Pass	Pass
Intensity difference between sample and a reference must be less than or equal to 0.5.		
IHC Staining	Pass	Pass
Background difference between sample and a reference must be less than or equal to 0.5.		
IHC - Tissue	Tissue Type	Brain
ELISA Titer	0.5 - 8.0 µg/ml	1.0 µg/ml
Bioburden	Pass	Pass
Visible bacterial colonies must be less than 100 CFU/mL.		
Source	Supernatant	Supernatant

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.



Product 760-4345
Batch V0004720



Bo Vang, Quality, QC Lab Technician II
Quality Assurance
Rocklin, California
US

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.

