



Certificate of Analysis

Product name

VENTANA Anti-E Cadherin (36)

Material No.

05905290001

Lot No.

E33587

Date of Manufacturing

2019-04-17

Expiry Date

2021-04-16

Parameter**Result****Unit**

Final Acceptance Testing

Pass

Ductal breast and lung carcinoma

The test batch must score 2.5+ or greater on ductal breast and lung carcinoma tissue

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result

The background intensity must be 0.5 or less

The intensity of the negative reagent control slide must be 0.5 or less

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

Issue date 2019-04-22

Page 1 of 2



Certificate of Analysis

Product name

VENTANA Anti-E Cadherin (36)

Material No.

05905290001

Lot No.

E33587

Date of Manufacturing

2019-04-17

Expiry Date

2021-04-16

This lot meets our specification

Disposition by

Lupe Villarreal

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2019-04-22

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>