

**CERTIFICATE OF ANALYSIS**


---

Product: Confirm™ anti-Progesterone Receptor (PR)( 1E2 ) Rabbit Monoclonal Primary Antibody

Ventana [REF]: 790-4296

Roche Number: 05278392001

Batch Number: F02791

Test	Acceptance Criteria	Result
Functional Staining on breast carcinoma tissue and reference tissue slide	Nuclear staining must be present in 51% or more of the tumor cells	PASS
	Background staining reading must be less than or equal to 0.5+ on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be less than or equal to 0.5+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2017-05-28

Manufacture date: 15-MAY-2015

Laughlin Hoskinson

*Laughlin Hoskinson 28-May-2015*

**Quality Department**

**Date**