



Certificate of Analysis

Product name

VENTANA ANTI-P63 (4A4)

Material No.

05867061001

Lot No.

N00128

Date of Manufacturing

2025-02-13

Expiry Date

2027-02-12

Parameter**Result****Unit**

Visual Inspection
Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Pass

Functional Staining

Pass

Tonsil and Prostate Carcinoma Tissues

For specific stain intensity, the test batch and reference batch must score 3.0 or greater in squamous epithelial cells of the tonsil tissue and 3.0 or greater in normal cells of the prostate carcinoma tissue.

The neoplastic cells of the prostate carcinoma tissue are used as an intra-tissue control, and specific stain intensity must be less than or equal to 0.5.

The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch for both tissues.

The background intensity must be 0.5 or less on both tissues for the test batch and reference batch.

The background intensity of the negative reagent control on both tissues must be 0.5 or less.

Bioburden

Pass

Issue date 2025-03-07

Page 1 of 2



Certificate of Analysis

Product name

VENTANA ANTI-P63 (4A4)

Material No.

05867061001

Lot No.

N00128

Date of Manufacturing

2025-02-13

Expiry Date

2027-02-12

Parameter**Result****Unit**

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Antonio Reyes

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-03-07

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>