

CERTIFICATE OF ANALYSIS


Product: Anti-Pan Keratin (AE1/AE3/PCK26) Primary Antibody
Ventana **REF**: 760-2135
Roche Number: 05266840001
Batch Number: F02377

Test	Acceptance Criteria	Results
Functional Staining on skin, liver, breast cancer, prostate cancer, and colon cancer. May use multi tissue block if available	Skin slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Liver slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Breast cancer slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Prostate cancer slide must score 3.5+ or greater on a scale of 0-4+	PASS
	Colon cancer slide must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be 0.5+ or less on a scale of 0-4+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2017-05-28

Manufacture date: 20-MAY-2015

Lucia Romero  03-JUN-2015
Quality Department **Date**