

CERTIFICATE OF ANALYSIS


Product Name: CONFIRM anti-Melanosome (HMB45) Mouse Monoclonal Primary Antibody
Ventana REF: 790-4366
Roche Number: 05479282001
Batch Number: D07679

Test	Acceptance Criteria	Result
Monoclonal positive melanoma tissue	Reference slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be less than or equal to 0.5+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2015-09-28

Manufacture date : 23-SEP-2013



Quality Department Date