

anti-CD8 (SP239) Rabbit Monoclonal Primary Antibody

REF 790-7176

09780041001

IVD  50




Figure 1. Anti-CD8 (SP239) antibody staining of T-cell lymphoma.

INTENDED USE

Anti-CD8 (SP239) Rabbit Monoclonal Primary Antibody is intended for laboratory use in the qualitative immunohistochemical detection of CD8 by light microscopy in sections of formalin-fixed, paraffin-embedded tissue stained on a BenchMark IHC/ISH instrument.

This product should be interpreted by a qualified pathologist in conjunction with histological examination, relevant

clinical information, and proper controls.

This antibody is intended for in vitro diagnostic (IVD) use.

SUMMARY AND EXPLANATION

T-cells are white blood cells that recognize foreign antigens, stimulate B-cell differentiation, and elicit immune-mediated cell death.^{1,2} As T-cells mature, they express and retract various surface markers such as cluster of differentiation (CD) molecules, allowing an immunophenotypic method to identify lymphocytes based on the markers they express.^{1,2} T-cell leukemias and lymphomas are heterogeneous lymphoproliferative disorders, resulting from the clonal expansion of a T-cell that originates from either the bone marrow or a secondary lymphoid structure, respectively.^{1,2}

CD8 is a 68 kDa heterodimeric, transmembrane glycoprotein expressed by cytotoxic T-cells and at low levels in NK-cells and a subpopulation of bone marrow cells.²⁻⁴ CD8 acts as a co-receptor for the T-cell receptor and interacts directly with major histocompatibility complex class I molecules on antigen-presenting cells, which stimulate T-cell activation.^{2,4} CD8 is absent on immature T-cells and expressed in later stages of T-cell development as the cell matures into a cytotoxic T-cell.^{1,5} CD8 is generally expressed in neoplasms derived from cytotoxic T-cells and is generally absent in lymphomas derived from helper T-cells.²⁻⁵ Some T-cell lymphomas, such as anaplastic large cell lymphoma, mycosis fungoides, and peripheral T-cell lymphoma may aberrantly express CD8.^{2,3,5}

The detection of CD8 by immunohistochemistry (IHC) with the anti-CD8 (SP239) Rabbit Monoclonal Primary Antibody (anti-CD8 (SP239) antibody) may be used to aid in the identification of a subset of normal T-cells and sub-classification of T-cell lymphoma.

PRINCIPLE OF THE PROCEDURE

Anti-CD8 (SP239) antibody binds to the CD8 glycoprotein in formalin-fixed, paraffin-embedded (FFPE) tissue sections. This antibody can be visualized using OptiView DAB IHC Detection Kit (Cat. No. 760-700 / 06396500001) or *ultraView* Universal DAB Detection Kit (Cat. No. 760-500 / 05269806001). Refer to the respective method sheet for further information.

MATERIAL PROVIDED

Anti-CD8 (SP239) antibody contains sufficient reagent for 50 tests.

One 5 mL dispenser of anti-CD8 (SP239) antibody contains approximately 51.0 µg of a rabbit monoclonal antibody.

The antibody is diluted in a phosphate buffer containing carrier protein and 0.10% ProClin 300, a preservative.

Specific antibody concentration is approximately 10.2 µg/mL. There is no known non-specific antibody reactivity observed in this product.

Anti-CD8 (SP239) antibody is a recombinant rabbit monoclonal antibody produced as purified cell culture supernatant.

Refer to the appropriate VENTANA detection kit method sheet for detailed descriptions of: Principle of the Procedure, Material and Methods, Specimen Collection and Preparation for Analysis, Quality Control Procedures, Troubleshooting, Interpretation of Results, and Limitations.

MATERIALS REQUIRED BUT NOT PROVIDED

Staining reagents, such as VENTANA detection kits and ancillary components, including negative and positive tissue control slides, are not provided.

Not all products listed in the method sheet may be available in all geographies. Consult your local support representative.

The following reagents and materials may be required for staining but are not provided:

1. Recommended control tissue
2. Microscope slides, positively charged
3. Rabbit Monoclonal Negative Control Ig (Cat. No. 790-4795 / 06683380001)
4. OptiView DAB IHC Detection Kit (Cat. No. 760-700 / 06396500001)
5. *ultraView* Universal DAB Detection Kit (Cat. No. 760-500 / 05269806001)
6. EZ Prep Concentrate (10X) (Cat. No. 950-102 / 05279771001)
7. Reaction Buffer Concentrate (10X) (Cat. No. 950-300 / 05353955001)
8. LCS (Predilute) (Cat. No. 650-010 / 05264839001)
9. ULTRA LCS (Predilute) (Cat. No. 650-210 / 05424534001)
10. Cell Conditioning Solution (CC1) (Cat. No. 950-124 / 05279801001)
11. ULTRA Cell Conditioning Solution (ULTRA CC1) (Cat. No. 950-224 / 05424569001)
12. Hematoxylin II (Cat. No. 790-2208 / 05277965001)
13. Bluing Reagent (Cat. No. 760-2037 / 05266769001)
14. General purpose laboratory equipment
15. BenchMark IHC/ISH instrument

STORAGE AND STABILITY

Upon receipt and when not in use, store at 2-8°C. Do not freeze.

To ensure proper reagent delivery and the stability of the antibody, replace the dispenser cap after every use and immediately place the dispenser in the refrigerator in an upright position.

Every antibody dispenser is expiration dated. When properly stored, the reagent is stable to the date indicated on the label. Do not use reagent beyond the expiration date.

SPECIMEN PREPARATION

Routinely processed, FFPE tissues are suitable for use with this primary antibody when used with VENTANA detection kits and BenchMark IHC/ISH instruments. The recommended tissue fixative is 10% neutral buffered formalin.⁶ Sections should be cut at approximately 4 µm in thickness and mounted on positively charged slides. Slides should be stained immediately, as antigenicity of cut tissue sections may diminish over time.

It is recommended that positive and negative controls be run simultaneously with unknown specimens.

WARNINGS AND PRECAUTIONS

1. For in vitro diagnostic (IVD) use.
2. For professional use only.
3. **CAUTION:** In the United States, Federal law restricts this device to sale by or on the order of a physician. (Rx Only)
4. Do not use beyond the specified number of tests.
5. ProClin 300 solution is used as a preservative in this reagent. It is classified as an irritant and may cause sensitization through skin contact. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Use protective clothing and gloves.
6. Positively charged slides may be susceptible to environmental stresses resulting in inappropriate staining. Ask your Roche representative for more information on how to use these types of slides.
7. Materials of human or animal origin should be handled as biohazardous materials and disposed of with proper precautions. In the event of exposure, the health directives of the responsible authorities should be followed.^{7,8}
8. Avoid contact of reagents with eyes and mucous membranes. If reagents come in contact with sensitive areas, wash with copious amounts of water.
9. Avoid microbial contamination of reagents as it may cause incorrect results.

10. For further information on the use of this device, refer to the BenchMark IHC/ISH instrument User Guide, and instructions for use of all necessary components located at dialog.roche.com.
11. Consult local and/or state authorities with regard to recommended method of disposal.
12. Product safety labeling primarily follows EU GHS guidance. Safety data sheet available for professional user on request.
13. To report suspected serious incidents related to this device, contact the local Roche representative and the competent authority of the Member State or Country in which the user is established.

This product contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

Table 1. Hazard information.

Hazard	Code	Statement
	H317	May cause an allergic skin reaction.
	P261	Avoid breathing mist or vapours.
	P272	Contaminated work clothing should not be allowed out of the workplace.
	P280	Wear protective gloves.
	P333 + P313	If skin irritation or rash occurs: Get medical advice/attention.
	P362 + P364	Take off contaminated clothing and wash it before reuse.
	P501	Dispose of contents/ container to an approved waste disposal plant.

This product contains CAS # 55965-84-9, a reaction mass of 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1)

STAINING PROCEDURE

VENTANA primary antibodies have been developed for use on BenchMark IHC/ISH instruments in combination with VENTANA detection kits and accessories. Refer to the tables below for recommended staining protocols.

This antibody has been optimized for specific incubation times but the user must validate results obtained with this reagent.

The parameters for the automated procedures can be displayed, printed and edited according to the procedure in the instrument User Guide. Refer to the appropriate VENTANA detection kit method sheet for more details regarding immunohistochemistry staining procedures.

For more details on the proper use of this device, refer to the inline dispenser method sheet associated with P/N 790-7176.

Table 2. Recommended staining protocol for anti-CD8 (SP239) antibody with OptiView DAB IHC Detection Kit on BenchMark IHC/ISH instruments.

Procedure Type	Method	
	GX	ULTRA or ULTRA PLUS ^a
Deparaffinization	Selected	
Cell Conditioning (Antigen Unmasking)	CC1, 64 minutes	ULTRA CC1, 64 minutes, 100 °C
Pre Primary Peroxidase Inhibitor	Selected	
Antibody (Primary)	12 minutes, 37 °C	12 minutes, 36 °C
OptiView HQ Linker	8 minutes (default)	
OptiView HRP Multimer	8 minutes (default)	

Procedure Type	Method	
	GX	ULTRA or ULTRA PLUS ^a
Counterstain	Hematoxylin II, 4 minutes	
Post Counterstain	Bluing, 4 minutes	

^a Concordance was demonstrated between BenchMark ULTRA and BenchMark ULTRA PLUS instruments using representative assays

Table 3. Recommended staining protocol for anti-CD8 (SP239) antibody with *ultraView* Universal DAB Detection Kit on BenchMark IHC/ISH instruments.

Procedure Type	Method	
	GX	ULTRA or ULTRA PLUS ^a
Deparaffinization	Selected	
Cell Conditioning (Antigen Unmasking)	CC1, Standard	ULTRA CC1, 64 minutes (Standard), 95 °C
Antibody (Primary)	32 minutes, 37 °C	32 minutes, 36 °C
Counterstain	Hematoxylin II, 4 minutes	
Post Counterstain	Bluing, 4 minutes	

^a Concordance was demonstrated between BenchMark ULTRA and BenchMark ULTRA PLUS instruments using representative assays

Due to variation in tissue fixation and processing, as well as general lab instrument and environmental conditions, it may be necessary to increase or decrease the primary antibody incubation, cell conditioning or protease pretreatment based on individual specimens, detection used, and reader preference. For further information on fixation variables, refer to "Immunohistochemistry Principles and Advances."⁹

NEGATIVE REAGENT CONTROL

In addition to staining with anti-CD8 (SP239) antibody, a second slide should be stained with the appropriate negative control reagent.

POSITIVE TISSUE CONTROL

A positive tissue control must be included with each staining run. Optimal laboratory practice is to include a positive control section on the same slide as the test tissue. This helps identify any failures applying reagents to the slide. Tissue with weak positive staining is best suited for quality control. Control tissue may contain both positive and negative staining elements and serve as both the positive and negative control. Control tissue should be fresh autopsy, biopsy, or surgical specimen, prepared or fixed as soon as possible in a manner identical to test sections.

Known positive tissue controls should be utilized only for monitoring the correct performance of reagents and instruments, not as an aid in determining specific diagnosis of test samples. If the positive tissue controls fail to demonstrate positive staining, results of the test specimen should be considered invalid.

Examples of positive control tissues for this antibody are normal tonsil and appendix.

STAINING INTERPRETATION / EXPECTED RESULTS

The cellular staining pattern for anti-CD8 (SP239) antibody is predominantly membranous and some granular cytoplasmic.

SPECIFIC LIMITATIONS

OptiView DAB IHC Detection Kit is generally more sensitive than *ultraView* Universal DAB Detection Kit. The user must validate the results obtained with this reagent and detection systems.

All assays might not be registered on every instrument. Please contact your local Roche representative for more information.

PERFORMANCE CHARACTERISTICS

ANALYTICAL PERFORMANCE

Staining tests for sensitivity, specificity, and precision were conducted and the results are listed below.

Sensitivity and Specificity

Table 4. Sensitivity/Specificity of anti-CD8 (SP239) antibody was determined by testing FFPE normal tissues.

Tissue	# positive / total cases	Tissue	# positive / total cases
Cerebrum ^a	0/3	Small intestine ^a	0/3
Cerebellum ^a	0/3	Colon ^a	0/5
Adrenal gland ^a	0/3	Colon (Congenital megacolon) ^a	0/1
Ovary ^a	0/3	Colon (Adenoma) ^a	0/1
Pancreas ^a	0/3	Colon (Tubular adenoma) ^a	0/1
Pituitary gland ^a	0/3	Liver ^a	0/3
Testis ^a	0/3	Salivary gland ^a	0/3
Thyroid ^a	0/3	Kidney ^a	0/3
Parathyroid ^a	0/3	Prostate ^a	0/3
Breast	0/3	Bladder ^a	0/3
Spleen	4/4	Uterus ^a	0/3
Tonsil	5/5	Placenta ^a	0/3
Thymus	3/3	Cervix ^a	0/3
Bone marrow	3/3	Skeletal muscle	0/3
Lung ^a	0/3	Skin	0/3
Heart ^a	0/3	Nerve	0/3
Esophagus ^a	0/3	Mesothelium	0/3
Stomach ^a	0/3	Lymph node	1/1

^a Cytotoxic/suppressor T-cells staining positive

Table 5. Sensitivity/Specificity of anti-CD8 (SP239) antibody was determined by testing a variety of FFPE neoplastic tissues.

Pathology	# positive / total cases
Glioblastoma (Cerebrum)	0/1
Meningioma (Cerebrum)	0/1
Oligodendroglioma (Cerebellum)	0/1
Ependymoma (Cerebellum)	0/1
Adenocarcinoma (Head and neck)	0/1
Squamous cell carcinoma (Head and neck)	0/1
Adenoma (Adrenal gland)	0/1
Pheochromocytoma (Adrenal gland)	0/1
Granulosa cell tumor (Ovary)	0/1
Serous carcinoma (Ovary)	0/1
Teratoma (Ovary)	0/1

Pathology	# positive / total cases
Islet cell tumor (Pancreas)	0/1
Pancreatic ductal adenocarcinoma (Pancreas)	0/1
Embryonal carcinoma (Testis)	0/1
Seminoma (Testis)	0/1
Follicular carcinoma (Thyroid)	0/1
Papillary carcinoma (Thyroid)	0/1
Ductal carcinoma in situ (DCIS) (Breast)	0/1
Invasive ductal carcinoma (Breast)	0/1
Invasive lobular carcinoma (Breast)	0/1
Adenocarcinoma (Lung)	0/1
Squamous cell carcinoma (Lung)	0/1
Small cell carcinoma (Lung)	0/1
Adenocarcinoma (Esophagus)	0/1
Squamous cell carcinoma (Esophagus)	0/1
Adenocarcinoma (Stomach)	0/1
Gastrointestinal stromal tumor (GIST) (Stomach)	0/1
Adenocarcinoma (Small intestine)	0/1
Gastrointestinal stromal tumor (GIST) (Small intestine)	0/1
Adenocarcinoma (Colon) ^a	0/87
Adenosquamous carcinoma (Colon) ^a	0/1
Mucinous adenocarcinoma (Colon)	0/8
Papillary adenocarcinoma (Colon)	0/3
Carcinoid (Appendix)	0/1
Cholangiocarcinoma (Liver)	0/1
Hepatocellular carcinoma (HCC) (Liver)	0/1
Pleomorphic adenoma (Salivary gland)	0/1
Warthin's tumor (Salivary gland)	0/1
Papillary renal adenoma (Kidney)	0/1
Renal cell carcinoma (Kidney)	0/1
Adenocarcinoma (Prostate)	0/2
Clear cell carcinoma (Uterus)	0/1
Endometrioid carcinoma (Uterus)	0/1
Leiomyoma (Uterus)	0/1
Leiomyosarcoma (Uterus)	0/1
Squamous cell carcinoma (Cervix)	0/1
Adenocarcinoma (Cervix)	0/1
Alveolar rhabdomyosarcoma (Muscle)	0/1
Myxoma (Muscle)	0/1
Basal cell carcinoma (Skin)	0/1
Invasive melanoma (Skin) ^a	0/1
Squamous cell carcinoma (Skin)	0/1
Neurofibrosarcoma (Peripheral nerve)	0/1

Pathology	# positive / total cases
Schwannoma (Peripheral nerve)	0/1
Anaplastic large cell lymphoma (ALCL)	0/5
Follicular lymphoma (Lymph node)	0/1
Hodgkin's lymphoma (Lymph node)	0/1
Diffuse large B-cell lymphoma	0/2
Peripheral T-cell lymphoma	20/44
NK/T-cell lymphoma	1/3
T-cell lymphoblastic lymphoma	0/4
T-Cell lymphoma, NOS	2/2
Squamous cell carcinoma (Bladder)	0/1
Urothelial carcinoma (Bladder)	0/1
Angiosarcoma (Soft tissue)	0/1
Liposarcoma (Soft tissue)	0/1
Multiple myeloma (Soft tissue)	0/1
Mesothelioma (Soft tissue)	0/1
Solitary fibrous tumor (Soft tissue)	0/1

^a Tumor Infiltrating Lymphocytes stain positive

Precision

Precision studies for anti-CD8 (SP239) antibody were completed to demonstrate:

- Between lot precision of the antibody.
- Within run and between day precision on a BenchMark ULTRA instrument.
- Between instrument precision on the BenchMark GX, BenchMark ULTRA / BenchMark ULTRA PLUS instrument.
- Between platform precision between the BenchMark GX, BenchMark ULTRA / BenchMark ULTRA PLUS instrument.

All studies met their acceptance criteria.

Precision on the BenchMark ULTRA PLUS instrument was demonstrated using representative assays. Studies included Within-run Repeatability, Between-day and Between-run Intermediate Precision. All studies met their acceptance criteria.

CLINICAL PERFORMANCE

Clinical performance data relevant to the intended purpose of anti-CD8 (SP239) antibody were assessed by systematic review of the literature. The data gathered support the use of the device in accordance with its intended purpose.

REFERENCES

1. Rich R, Fleisher T, Shearer W, Frew A, Weyand C. Clinical Immunology Principles and Practice, 5th edition. Vol 5. Amsterdam, Netherlands: Elsevier; 2018.
2. Dabbs DJ. Diagnostic Immunohistochemistry Theranostic and Genomic Applications, 5th edition. Vol 5. Amsterdam, Netherlands: Elsevier; 2019.
3. Higgins RA, Blankenship JE, Kinney MC. Application of immunohistochemistry in the diagnosis of non-Hodgkin and Hodgkin lymphoma. Arch Pathol Lab Med. 2008;132(3):441-461.
4. Naeim F. Principles of Immunophenotyping. In: Naeim F, Rao PN, Grody WW, eds. Hematopathology: Morphology, Immunophenotype, Cytogenetics, and Molecular Approaches. Cambridge, MA: Academic Press; 2009.
5. Swerdlow SH, Campo E, Harris NL, et al. WHO Classification of Tumours of Haematopoietic and Lymphoid Tissues, 4th edition. Vol 4. Lyon, France: International Agency for Research on Cancer; 2008.
6. Carson F, Hladik C. Histotechnology: A Self Instructional Text, 3rd edition. Hong Kong: American Society for Clinical Pathology Press; 2009.
7. Occupational Safety and Health Standards: Occupational exposure to hazardous chemicals in laboratories. (29 CFR Part 1910.1450). Fed. Register.

8. Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
9. Roche PC, Hsi ED. Immunohistochemistry-Principles and Advances. Manual of Clinical Laboratory Immunology, 6th edition. In: NR Rose, ed. ASM Press; 2002.

NOTE: A point (period/stop) is always used in this document as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

The summary of safety and performance can be found here:

<https://ec.europa.eu/tools/eudamed>

Symbols

Ventana uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.rocke.com for definition of symbols used):



Global Trade Item Number



Unique Device Identification



Indicates the entity importing the medical device into the European Union

REVISION HISTORY

Rev	Updates
A	Initial Release

INTELLECTUAL PROPERTY

VENTANA, BENCHMARK, OPTIVIEW, *ultraView*, and the VENTANA logo are trademarks of Roche. All other trademarks are the property of their respective owners.

© 2023 Ventana Medical Systems, Inc.

CONTACT INFORMATION



Ventana Medical Systems, Inc.
1910 E. Innovation Park Drive
Tucson, Arizona 85755
USA
+1 520 887 2155
+1 800 227 2155 (USA)

www.rocke.com



Roche Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim
Germany
+800 5505 6606

