

## Customer Address:

Customer Number : 2049990013  
 Customer Mat # : 05269784001  
 Roche Diagnostics Corporation  
 Accounts Payable  
 Po Box 50457  
 INDIANAPOLIS IN 46250-0417  
 UNITED STATES

## Certificate of Analysis

**Product Name :** Glial Fibrillary Acidic Protein (EP672Y) Rabbit Monoclonal Antibody  
**Product Number :** 760-4345  
**Batch Number :** V0002483  
 Source Batch : 0000074186  
 Storage Temperature : 2°C - 8°C  
 Expiry Date : 2022-08-31  
 Quality Release Date : 2021-02-23  
 Date of Manufacture : 2019-09-23

| Test  | Specification                    | Result                           |
|---|----------------------------------|----------------------------------|
| Appearance (Turbidity)  | Free of Visible Foreign Material | Free of Visible Foreign Material |
| Appearance (Color)  | Green Transparent                | Green Transparent                |
| Appearance (Form)   | Liquid                           | Liquid                           |
| IHC Staining<br>Intensity difference between sample and a reference must be less than or equal to 0.5.  | Pass                             | Pass                             |
| IHC Staining<br>Background difference between sample and a reference must be less than or equal to 0.5. | Pass                             | Pass                             |
| IHC - Tissue  | Tissue Type                      | Brain                            |
| ELISA Titer   | 1.00 - 10.00 µg/ml               | 1.07 µg/ml                       |
| Bioburden<br>Visible bacterial colonies must be less than 100 CFU/mL.                                   | Pass                             | Pass                             |
| Source  | Supernatant                      | Supernatant                      |

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.



**Product** 760-4345  
**Batch** V0002483

Patrick Strather

Patrick Strather, QC Inspector  
Quality Assurance  
Rocklin, California  
US

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.

