

Customer Address:

Customer Number : 2049990013
 Customer Mat # : 08313440001
 Roche Diagnostics Corporation
 Accounts Payable
 Po Box 50457
 INDIANAPOLIS IN 46250-0417
 UNITED STATES

Certificate of Analysis

Product Name : Cytokeratin (OSCAR) Mouse Monoclonal Antibody
Product Number : 760-6078
Batch Number : V0003902
 Source Batch : 0000229508
 Storage Temperature : 2°C - 8°C
 Expiry Date : 2026-03-31
 Quality Release Date : 2023-05-31
 Date of Manufacture : 2023-04-14

Test	Specification	Result
Appearance (Turbidity)	Free of Visible Foreign Material	Free of Visible Foreign Material
Appearance (Color)	Green Transparent	Green Transparent
Appearance (Form)	Liquid	Liquid
IHC Staining	Pass	Pass
Intensity difference between sample and a reference must be less than or equal to 0.5.		
IHC Staining	Pass	Pass
Background difference between sample and a reference must be less than or equal to 0.5.		
IHC - Tissue	Tissue Type	Prostate
ELISA Titer	0.10 - 1.00 µg/ml	0.17 µg/ml
Bioburden	Pass	Pass
Visible bacterial colonies must be less than 100 CFU/mL.		
Source	Supernatant	Supernatant

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.



Product 760-6078
Batch V0003902



JUNIOR NUNEZ, Quality, QC Lab Technician II

Quality Assurance

Rocklin, California

US

Cell Marque™ (a Sigma-Aldrich Company) warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. For further inquiries, please contact Technical/Customer Service. Purchase must determine the suitability of the product for its particular use. See website or packing slip for additional terms and conditions of sale.

