



Certificate of Analysis

Product name

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

Material No.

05278368001

Lot No.

J20324

Date of Manufacturing

2022-08-05

Expiry Date

2024-02-03

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Functional Staining # iVIEW DAB Kit

Pass

Breast Carcinoma Tissue

Gastric Carcinoma Tissue

Qualified 4 in 1 Cell Line Controls

The specific staining of the test batch must be acceptable on all tissues and cell line cores using the iVIEW DAB IHC.

The test batch background intensity must not exceed 0.5 pt. on any tissue or cell line core.

All tissues and cell line cores stained with the negative control reagent must be negative for specific staining.

Functional Staining # ultraVIEW DAB Kit

Pass

Breast Carcinoma Tissue

Gastric Carcinoma Tissue

Qualified 4 in 1 Cell Line Controls

Issue date 2022-09-09

Page 1 of 2



Certificate of Analysis

Product name

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

Material No.

05278368001

Lot No.

J20324

Date of Manufacturing

2022-08-05

Expiry Date

2024-02-03

Parameter**Result****Unit**

The specific staining of the test batch must be acceptable on all tissues and cell line cores using the ultraVIEW DAB IHC.

The test batch background intensity must not exceed 0.5 pt. on any tissue or cell line core.

All tissues and cell line cores stained with the negative control reagent must be negative for specific staining.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Lupe Villarreal

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2022-09-09

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>