



Certificate of Analysis

Product name

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

Material No.

05278368001

Lot No.

F10545

Date of Manufacturing

2019-05-20

Expiry Date

2020-11-15

Parameter**Result****Unit**

Functional Staining

Pass

Breast Carcinoma Tissue, Gastric Carcinoma Tissue,
and 4 in 1 Cell Line.The specific staining of the test batch must be
acceptable on all tissues and cell line cores using
both the iView DAB IHC and the ultraView DAB IHC
detection kits.The test batch background intensity must not exceed
0.5 pt on any tissue or cell line core.All tissues and cell line cores stained with the negative
control reagent must be negative for specific staining.

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and

Issue date 2019-06-12

Page 1 of 2



Certificate of Analysis

Product name

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

Material No.

05278368001

Lot No.

F10545

Date of Manufacturing

2019-05-20

Expiry Date

2020-11-15

Parameter**Result****Unit**

substantially free of intrinsic particulate matter.

This lot meets our specification

Disposition by

Lupe Villarreal

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2019-06-12

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>